

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3390 **SAS No.:** P3390 **SDG NO.:** P3390

Continuing Calib Date: 07/30/2024 **Initial Calibration Date(s):** 07/30/2024 07/30/2024

Continuing Calib Time: 13:04 **Initial Calibration Time(s):** 10:41 11:19

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	6.63	6.63	6.53	6.73	0.00
EDB	2.87	2.87	2.77	2.97	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3390 **SAS No.:** P3390 **SDG NO.:** P3390

Continuing Calib Date: 07/30/2024 **Initial Calibration Date(s):** 07/30/2024 07/30/2024

Continuing Calib Time: 13:04 **Initial Calibration Time(s):** 10:41 11:19

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.61	5.61	5.51	5.71	0.00
EDB	2.26	2.26	2.16	2.36	0.00