

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/30/2024 07/30/2024

Client Sample No.: CCAL01 Date Analyzed: 07/30/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR067878.D Time Analyzed: 11:56

| COMPOUND | RT | RT WINDOW | | CALC AMOUNT(ng) | NOM AMOUNT(ng) | %D |
|----------|-------|-----------|-------|-----------------|----------------|-----|
| | | FROM | TO | | | |
| DBCP | 6.628 | 6.528 | 6.728 | 0.110 | 0.100 | 0.0 |
| EDB | 2.871 | 2.770 | 2.970 | 0.090 | 0.100 | 0.0 |

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/30/2024 07/30/2024

Client Sample No.: CCAL01 Date Analyzed: 07/30/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR067878.D Time Analyzed: 11:56

| COMPOUND | RT | RT WINDOW | | CALC AMOUNT(ng) | NOM AMOUNT(ng) | %D |
|----------|-------|-----------|-------|-----------------|----------------|-----|
| | | FROM | TO | | | |
| DBCP | 5.606 | 5.506 | 5.706 | 0.100 | 0.100 | 0.0 |
| EDB | 2.257 | 2.157 | 2.357 | 0.100 | 0.100 | 0.0 |