

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3390 **SAS No.:** P3390 **SDG NO.:** P3390

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/30/2024 07/30/2024

Client Sample No.: CCAL02 **Date Analyzed:** 07/30/2024

Lab Sample No.: M8011-504 CCC 0. **Data File :** PR067883.D **Time Analyzed:** 13:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.627	6.528	6.728	0.110	0.100	0.0
EDB	2.869	2.770	2.970	0.090	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/30/2024 07/30/2024

Client Sample No.: CCAL02 Date Analyzed: 07/30/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR067883.D Time Analyzed: 13:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	5.606	5.506	5.706	0.100	0.100	0.0
EDB	2.256	2.157	2.357	0.100	0.100	0.0