

Manual Integration Report

Sequence:	PR073024	Instrument	ECD_r
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
M8011.504.1 0.025 PPB ICC	PR067875.D	DBCP	Ankita	7/30/2024 4:25:32 PM	Sohil	7/30/2024 4:25:53	Peak Integrated by Software incorrectly
M8011.504.1 0.025 PPB ICC	PR067875.D	EDB	Ankita	7/30/2024 4:25:32 PM	Sohil	7/30/2024 4:25:53	Peak Integrated by Software incorrectly
M8011.504.1 0.025 PPB ICC	PR067875.D	EDB #2	Ankita	7/30/2024 4:25:32 PM	Sohil	7/30/2024 4:25:53	Peak Integrated by Software incorrectly
P3390-08	PR067881.D	DBCP	Ankita	7/30/2024 4:25:35 PM	Sohil	7/30/2024 4:26:01	Peak Integrated by Software incorrectly