

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR073024

Review By	Ankita	Review On	7/30/2024 4:25:06 PM
Supervise By	Sohil	Supervise On	7/30/2024 4:26:18 PM
SubDirectory	PR073024	HP Acquire Method	HP Processing Method PR073024
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23554,PP23555,PP23556,PP23557,PP23558		
CCC	PP23561		
Internal Standard/PEM			
ICV/I.BLK	PP23559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PR067870.D	30 Jul 2024 10:31		AJ\MA	Ok
2	M8011.504.1 0.5 PPB I	M8011.504.1 0.5 PPB I	PR067871.D	30 Jul 2024 10:41		AJ\MA	Ok
3	M8011.504.1 0.25 PPB	M8011.504.1 0.25 PPB	PR067872.D	30 Jul 2024 10:50		AJ\MA	Ok
4	M8011.504.1 0.1 PPB I	M8011.504.1 0.1 PPB I	PR067873.D	30 Jul 2024 11:00		AJ\MA	Ok
5	M8011.504.1 0.05 PPB	M8011.504.1 0.05 PPB	PR067874.D	30 Jul 2024 11:10		AJ\MA	Ok
6	M8011.504.1 0.025 PPB	M8011.504.1 0.025 PPB	PR067875.D	30 Jul 2024 11:19		AJ\MA	Ok,M
7	M8011.504.1 0.1 PPB I	ICVPR073024	PR067876.D	30 Jul 2024 11:33		AJ\MA	Ok
8	RT CHECK	RT CHECK	PR067877.D	30 Jul 2024 11:47		AJ\MA	Ok
9	M8011-504 CCC 0.1 PP	M8011-504 CCC 0.1 PP	PR067878.D	30 Jul 2024 11:56		AJ\MA	Ok
10	Low-Level LFB-MDL Ch	Low-Level LFB-MDL Ch	PR067879.D	30 Jul 2024 12:25		AJ\MA	Ok,M
11	PB162361BL	PB162361BL	PR067880.D	30 Jul 2024 12:35		AJ\MA	Ok
12	P3390-08	LOQ-WATER-02-QT3-2	PR067881.D	30 Jul 2024 12:45		AJ\MA	Ok,M
13	P3390-09	MDL-WATER-03-QT3-2	PR067882.D	30 Jul 2024 12:54		AJ\MA	Ok,M
14	M8011-504 CCC 0.1 PP	M8011-504 CCC 0.1 PP	PR067883.D	30 Jul 2024 13:04		AJ\MA	Ok
15	PB162360BL	PB162360BL	PR067884.D	30 Jul 2024 13:14		AJ\MA	Ok
16	PB162360BS	PB162360BS	PR067885.D	30 Jul 2024 13:23		AJ\MA	Ok
17	PB162360BSD	PB162360BSD	PR067886.D	30 Jul 2024 13:33		AJ\MA	Ok,M
18	P3204-11	PT-EDBCP-WS	PR067887.D	30 Jul 2024 13:43		AJ\MA	Dilution

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR073024

Review By	Ankita	Review On	7/30/2024 4:25:06 PM
Supervise By	Sohil	Supervise On	7/30/2024 4:26:18 PM
SubDirectory	PR073024	HP Acquire Method	HP Processing Method PR073024

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP23554,PP23555,PP23556,PP23557,PP23558
CCC Internal Standard/PEM	PP23561
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23559

19	P3204-11DL	PT-EDBCP-WSDL	PR067888.D	30 Jul 2024 13:56	Need more dil	AJ\MA	Not Ok
20	P3204-11DL	PT-EDBCP-WSDL	PR067889.D	30 Jul 2024 14:17		AJ\MA	Ok
21	M8011-504 CCC 0.1 PF	M8011-504 CCC 0.1 PF	PR067890.D	30 Jul 2024 14:29		AJ\MA	Ok

M : Manual Integration