



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger Site Princeton NJ

Project # N/A

Chemtech Project # P3426

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 07/31/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Hexavalent Chromium, Metals Group4, SVOC-SIMGroup1, SVOCMS Group3, SVOCMS Group6, VOCMS Group3 and VOCMS Group6. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for 927-K1-WS-073124-FD [2-Methylnaphthalene-d10 - 16%, Fluoranthene-d10 - 22%] and 927-K1-WS-073124-FDRE [2-Methylnaphthalene-d10 - 17%, Fluoranthene-d10 - 22%] these compounds did not meet the NJDKQP criteria and in-house criteria, All the failure samples in surrogates were reanalyzed to confirm the results as per method and reported in the data and 927-K1-WS-073124-FD [2-Fluorobiphenyl - 21%, Nitrobenzene-d5 - 21%] and 927-K1-WS-073124-FDRE [2-Fluorobiphenyl - 20%, Nitrobenzene-d5 - 21%] these compounds did not meet the NJDKQP criteria but met the in-house criteria.

The Internal Standards Areas met the acceptable requirements except for 927-K1-WS-073124-FD, 927-K1-WS-073124-FDRE , All the failure samples in Internal Standard were reanalyzed to confirm the results as per method and reported in the data.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD met criteria .

The Blank Spike met requirements for all samples .



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements .

The Continuous Calibration File ID BN033212.D met the requirements except for Benzo(b)fluoranthene is failing marginally low and 2,4,6-Tribromophenol and Phenol-d6 , failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken.

The Continuous Calibration File ID BN033228.D met the requirements except for Benzo(b)fluoranthene is failing marginally low and 2,4,6-Tribromophenol failure surrogate is not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

Less volume was taken for sample # 927-K1-WS-073124-FD at the extraction due to Limited volume received.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____