



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080724\
Data File : BF138833.D
Acq On : 07 Aug 2024 10:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jul 30 17:50:01 2024






AutoFind: Scans 1617, 1618, 1619; Background Corrected with Scan 1611

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.0	27181	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.2	23713	PASS
70	69	0.00	2	0.6	146	PASS
127	198	10	80	48.0	29053	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	60469	PASS
199	198	5	9	6.4	3864	PASS
275	198	10	60	26.2	15814	PASS
365	198	1	100	3.3	2002	PASS
441	198	0.01	100	13.5	8148	PASS
442	442	50	100	100.0	51789	PASS
443	442	15	24	18.6	9641	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080724\
Data File : BF138833.D
Acq On : 07 Aug 2024 10:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 07 11:59:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration



TIC: BF138833.D\data.ms

(70) Pentachlorophenol (C)

11.169min (-0.006) 57532.53 ng




response 68125

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.50	62.36
264.00	62.20	63.20
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080724\
Data File : BF138833.D
Acq On : 07 Aug 2024 10:30
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 07 11:59:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:50:01 2024
Response via : Initial Calibration



TIC: BF138833.D\data.ms

(77) Benzidine

12.698min (-0.006) 0.00 ng

response 323216

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	14.17
183.00	12.00	12.73
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/7/2024	BNA_F	<u>BF138833.D</u>
Compound Name	Response	Retention Time
DDT	163406	13.516
DDD	2660	13.245
DDE	908	12.88
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3568	166974	2.14

Instrument :
BNA_F
ClientSampleId :
DFTPP