

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3426 SAS No.: P3426 SDG NO.: P3426

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/07/2024

Lab File ID: BF138834.D Time Analyzed: 11:00

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	53283	6.84	211206	8.12	116265	9.88
UPPER LIMIT	106566	7.34	422412	8.622	232530	10.381
LOWER LIMIT	26641.5	6.34	105603	7.622	58132.5	9.381
EPA SAMPLE NO.						
01 927-K1-WS-073124	42119	6.84	170418	8.12	92774	9.87
02 927-K1-WS-073124-FD	84689	6.84	334542	8.12	183422	9.88

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.