

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3426 SDG NO.: P3426

Lab File ID: BF138833.D

DFTPP Injection Date: 08/07/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:30

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	45
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	39.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	48
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	13.5
442	Greater than 50% of mass 198	85.6
443	15.0 - 24.0% of mass 442	15.9 (18.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF138834.D	08/07/2024	11:00
927-K1-WS-073124	P3426-01	BF138845.D	08/07/2024	16:35
927-K1-WS-073124-FD	P3426-02	BF138846.D	08/07/2024	17:05