

Surrogate Summary

SDG No.: **P3426**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P3426-01	927-K1-WS-073124	1,2-Dichloroethane-d4	50	56.8	114	70 (74)	130 (125)
		Dibromofluoromethane	50	52.6	105	70 (75)	130 (124)
		Toluene-d8	50	53.5	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.8	114	70 (77)	130 (121)
P3426-02	927-K1-WS-073124-FD	1,2-Dichloroethane-d4	50	56.4	113	70 (74)	130 (125)
		Dibromofluoromethane	50	53.2	106	70 (75)	130 (124)
		Toluene-d8	50	53.5	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.3	113	70 (77)	130 (121)
VN0812WBL01	VN0812WBL01	1,2-Dichloroethane-d4	50	56.6	113	70 (74)	130 (125)
		Dibromofluoromethane	50	53.6	107	70 (75)	130 (124)
		Toluene-d8	50	53.9	108	70 (86)	130 (113)
		4-Bromofluorobenzene	50	59.1	118	70 (77)	130 (121)
VN0812WBS01	VN0812WBS01	1,2-Dichloroethane-d4	50	54.1	108	70 (74)	130 (125)
		Dibromofluoromethane	50	52.1	104	70 (75)	130 (124)
		Toluene-d8	50	51.9	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
VN0812WBSD0	VN0812WBSD01	1,2-Dichloroethane-d4	50	48.8	98	70 (74)	130 (125)
		Dibromofluoromethane	50	46.0	92	70 (75)	130 (124)
		Toluene-d8	50	46.4	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.3	91	70 (77)	130 (121)

