

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3426 SAS No.: P3426 SDG NO.: P3426
 Lab File ID: VN083225.D Date Analyzed: 08/12/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	128278	13.788				
UPPER LIMIT	256556	14.288				
LOWER LIMIT	64139	13.288				
EPA SAMPLE NO.						
927-K1-WS-073124	118509	13.79				
927-K1-WS-073124-FD	115072	13.79				
VN0812WBL01	137198	13.79				
VN0812WBS01	130523	13.79				
VN0812WBSD01	129761	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.