

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: P3426 SAS No.: P3426 SDG No.: P3426

Instrument ID: MSVOA\_N Calibration Date/Time: 08/12/2024 09:48

Lab File ID: VN083225.D Init. Calib. Date(s): 08/07/2024 08/07/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:33 12:34

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|--------------------------------|-------|--------|------------|-------|-------|
| Dichlorodifluoromethane        | 0.567 | 0.591  |            | 4.23  | 20    |
| Chloromethane                  | 0.581 | 0.583  | 0.1        | 0.34  | 20    |
| Vinyl Chloride                 | 0.592 | 0.608  |            | 2.7   | 20    |
| Bromomethane                   | 0.368 | 0.385  |            | 4.62  | 20    |
| Chloroethane                   | 0.371 | 0.384  |            | 3.5   | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.539 | 0.582  |            | 7.98  | 20    |
| 1,1-Dichloroethene             | 0.554 | 0.542  |            | -2.17 | 20    |
| Acetone                        | 0.278 | 0.307  |            | 10.43 | 20    |
| Carbon Disulfide               | 1.622 | 1.491  |            | -8.08 | 20    |
| Methyl tert-butyl Ether        | 2.001 | 2.059  |            | 2.9   | 20    |
| Methylene Chloride             | 0.641 | 0.615  |            | -4.06 | 20    |
| trans-1,2-Dichloroethene       | 0.573 | 0.569  |            | -0.7  | 20    |
| 1,1-Dichloroethane             | 1.073 | 1.130  | 0.1        | 5.31  | 20    |
| Cyclohexane                    | 1.055 | 0.991  |            | -6.07 | 20    |
| 2-Butanone                     | 0.428 | 0.411  |            | -3.97 | 20    |
| Carbon Tetrachloride           | 0.532 | 0.586  |            | 10.15 | 20    |
| cis-1,2-Dichloroethene         | 0.691 | 0.707  |            | 2.32  | 20    |
| Chloroform                     | 1.115 | 1.206  |            | 8.16  | 20    |
| 1,1,1-Trichloroethane          | 1.055 | 1.119  |            | 6.07  | 20    |
| Methylcyclohexane              | 0.580 | 0.609  |            | 5     | 20    |
| Benzene                        | 1.406 | 1.496  |            | 6.4   | 20    |
| 1,2-Dichloroethane             | 0.512 | 0.557  |            | 8.79  | 20    |
| Trichloroethene                | 0.335 | 0.356  |            | 6.27  | 20    |
| Bromodichloromethane           | 0.537 | 0.580  |            | 8.01  | 20    |
| Toluene                        | 0.889 | 0.963  |            | 8.32  | 20    |
| 1,1,2-Trichloroethane          | 0.318 | 0.347  |            | 9.12  | 20    |
| Dibromochloromethane           | 0.385 | 0.425  |            | 10.39 | 20    |
| Tetrachloroethene              | 0.331 | 0.353  |            | 6.65  | 20    |
| Chlorobenzene                  | 1.105 | 1.171  | 0.3        | 5.97  | 20    |
| Ethyl Benzene                  | 2.027 | 2.160  |            | 6.56  | 20    |
| m/p-Xylenes                    | 0.759 | 0.814  |            | 7.25  | 20    |
| o-Xylene                       | 0.749 | 0.784  |            | 4.67  | 20    |
| Isopropylbenzene               | 4.182 | 4.339  |            | 3.75  | 20    |
| 1,4-Dichlorobenzene            | 1.762 | 1.825  |            | 3.58  | 20    |
| 1,2-Dichlorobenzene            | 1.692 | 1.726  |            | 2.01  | 20    |
| 1,2-Dichloroethane-d4          | 0.712 | 0.796  |            | 11.8  | 20    |
| Dibromofluoromethane           | 0.312 | 0.343  |            | 9.94  | 20    |
| Toluene-d8                     | 1.164 | 1.296  |            | 11.34 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: P3426 SAS No.: P3426 SDG No.: P3426

Instrument ID: MSVOA\_N Calibration Date/Time: 08/12/2024 09:48

Lab File ID: VN083225.D Init. Calib. Date(s): 08/07/2024 08/07/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:33 12:34

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND             | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|----------------------|-------|--------|------------|-------|-------|
| 4-Bromofluorobenzene | 0.454 | 0.507  |            | 11.67 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.