

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3426 SAS No.: P3426 SDG NO.: P3426
 Lab File ID: VN083225.D Date Analyzed: 08/12/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	179688	8.22	304896	9.10	267164	11.87
UPPER LIMIT	359376	8.724	609792	9.6	534328	12.365
LOWER LIMIT	89844	7.724	152448	8.6	133582	11.365
EPA SAMPLE NO.						
927-K1-WS-073124	133768	8.22	260137	9.10	268551	11.87
927-K1-WS-073124-FD	135406	8.22	259109	9.10	264619	11.87
VN0812WBL01	155185	8.22	297997	9.10	309712	11.87
VN0812WBS01	193104	8.22	331101	9.10	283776	11.87
VN0812WBSD01	186747	8.22	319813	9.10	275661	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.