

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger Site Princeton NJ

Project # N/A

Chemtech Project # P3429

Test Name: SVOCMS Group6

A. Number of Samples and Date of Receipt:

4 Water samples were received on 07/31/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Hexavalent Chromium, Mercury, Metals Group4, SVOC-SIMGroup1, SVOCMS Group3, SVOCMS Group6 and VOCMS Group6. This data package contains results for SVOCMS Group6.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um dfThe analysis of SVOCMS Group6 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for MLS-15-70-85MSD [2,4,6-Tribromophenol - 115%], PB162423BL [2,4 and6-Tribromophenol - 120%] these compounds did not meet the NJDKQP criteria but met the in-house criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {P3415-04MS} with File ID: BF138909.D recoveries met the requirements for all compounds except for Benzo(k)fluoranthene[131%] this compound did not meet the NJDKQP criteria but met the in-house criteria and Benzaldehyde[0%] this compound did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The MSD {P3415-05MSD} with File ID: BF138910.D recoveries met the acceptable requirements except for Acenaphthylene[131%],Benzo(a)pyrene[133%], Benzo(k)fluoranthene[140%], Chrysene[132%] and Di-n-butylphthalate[137%] these compounds did not meet the NJDKQP criteria but met the in-house criteria and Benzaldehyde[8%] this compound did not meet the NJDKQP criteria and in-house criteria due to matrix interference.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The RPD for {P3415-05MSD} with File ID: BF138910.D met criteria except for Benzaldehyde[200%] due to difference in results of MS and MSD.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID BF138834.D met the requirements except for Pentachlorophenol but no positive hits in associated samples therefore no corrective action taken.

The Continuous Calibration File ID BF138879.D met the requirements except for Benzaldehyde is failing marginally low therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____