

LB131960

7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

41267

A= 0.0000e+000

B= 2.4469e-004

C= -5.8539e-002

Rho= 0.9990596

Accept =Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.028	-0.028	123	0.000	123	0				1
0.2	0.200	0.194	-0.006	1032	0.0 %	1032					-3
2.5	2.500	2.456	-0.044	10275	0.0 %	10275					-2
5.0	5.000	5.292	0.292	21867	0.0 %	21867					6
7.5	7.500	7.247	-0.253	29857	0.0 %	29857					3
10.0	10.000	10.039	0.039	41267	0.0 %	41267					0

LB131960 INSTRUMKENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Aug 2024 08:29:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	123	-	0.0000	7470A	PPB	13 Aug 2024 08:36:49	S	Std
0.2 - 1	50.2	1032	-	0.2000	7470A	PPB	13 Aug 2024 08:39:05	S	Std
2.5 - 1	52.5	10275	-	2.5000	7470A	PPB	13 Aug 2024 08:41:23	S	Std
5.0 - 1	55	21867	-	5.0000	7470A	PPB	13 Aug 2024 08:43:40	S	Std
7.5 - 1	57.5	29857	-	7.5000	7470A	PPB	13 Aug 2024 08:45:57	S	Std
10.0 - 1	510	41267	-	10.0000	7470A	PPB	13 Aug 2024 08:50:50	S	Std
ICV72 - 1	ICV72	16618	4.0078	-	7470A	PPB	13 Aug 2024 08:53:41	U	SMPL
ICB72 - 1	ICB72	-81	-0.0784	-	7470A	PPB	13 Aug 2024 08:55:55	U	SMPL
CCV35 - 1	CCV35	18763	4.5327	-	7470A	PPB	13 Aug 2024 08:58:12	U	SMPL
CCB35 - 1	CCB35	-64	-0.0742	-	7470A	PPB	13 Aug 2024 09:03:02	U	SMPL
CRA - 1	CRA	1203	0.2358	-	7470A	PPB	13 Aug 2024 09:10:24	U	SMPL
HighStd - 1	HighStd	41121	10.0035	-	7470A	PPB	13 Aug 2024 09:12:39	U	SMPL
ChkStd - 1	ChkStd	27661	6.7100	-	7470A	PPB	13 Aug 2024 09:14:54	U	SMPL
PB162665BL - 1	PBW	-169	-0.0999	-	7470A	PPB	13 Aug 2024 09:17:12	U	SMPL
PB162665BS - 1	LCSW	16637	4.0124	-	7470A	PPB	13 Aug 2024 09:19:29	U	SMPL
P3526-02 - 1	SP-1-COMP	-178	-0.1021	-	7470A	PPB	13 Aug 2024 09:21:45	U	SMPL
P3534-01 - 1	PIPE-BLAST-GRIT	6	-0.0571	-	7470A	PPB	13 Aug 2024 09:24:03	U	SMPL
P3534-03 - 1	HP-IP	24	-0.0527	-	7470A	PPB	13 Aug 2024 09:26:20	U	SMPL
P3534-03DUP - 1	HP-IPDUP	31	-0.0510	-	7470A	PPB	13 Aug 2024 09:28:37	U	SMPL
P3534-03MS - 1	HP-IPMS	19735	4.7705	-	7470A	PPB	13 Aug 2024 09:30:53	U	SMPL
CCV36 - 1	CCV36	19728	4.7688	-	7470A	PPB	13 Aug 2024 09:33:08	U	SMPL
CCB36 - 1	CCB36	-104	-0.0840	-	7470A	PPB	13 Aug 2024 09:35:25	U	SMPL
P3534-03MSD - 1	HP-IPMSD	15594	3.7572	-	7470A	PPB	13 Aug 2024 09:37:42	U	SMPL
PB162681BL - 1	PBW	-62	-0.0737	-	7470A	PPB	13 Aug 2024 09:39:57	U	SMPL
PB162681BS - 1	LCSW	16751	4.0403	-	7470A	PPB	13 Aug 2024 09:42:14	U	SMPL
P3530-02 - 1	WC-A-04-202408	-188	-0.1045	-	7470A	PPB	13 Aug 2024 09:44:29	U	SMPL
P3530-02DUP - 1	WC-A-04-202408DUP	-28	-0.0654	-	7470A	PPB	13 Aug 2024 09:46:47	U	SMPL
P3530-02MS - 1	WC-A-04-202408MS	14987	3.6087	-	7470A	PPB	13 Aug 2024 09:49:03	U	SMPL
P3530-02MSD - 1	WC-A-04-202408MSD	13826	3.3246	-	7470A	PPB	13 Aug 2024 09:51:19	U	SMPL
P3530-03 - 1	WC-A-05-202408	-95	-0.0818	-	7470A	PPB	13 Aug 2024 09:53:38	U	SMPL
P3534-05 - 1	PIPE-BLAST-GRIT(FB)	11	-0.0558	-	7470A	PPB	13 Aug 2024 09:55:57	U	SMPL
P3534-06 - 1	HP-IP(FB)	32	-0.0507	-	7470A	PPB	13 Aug 2024 09:58:13	U	SMPL
CCV37 - 1	CCV37	20451	4.9457	-	7470A	PPB	13 Aug 2024 10:00:28	U	SMPL
CCB37 - 1	CCB37	14	-0.0551	-	7470A	PPB	13 Aug 2024 10:05:27	U	SMPL
PB162680BL - 1	PBW	120	-0.0292	-	7470A	PPB	13 Aug 2024 10:07:43	U	SMPL
PB162680BS - 1	LCSW	15043	3.6224	-	7470A	PPB	13 Aug 2024 10:09:58	U	SMPL
P3426-01 - 1	927-K1-WS-073124	-94	-0.0815	-	7470A	PPB	13 Aug 2024 10:12:13	U	SMPL
P3426-02 - 1	927-K1-WS-073124-FD	31	-0.0510	-	7470A	PPB	13 Aug 2024 10:14:30	U	SMPL
P3429-01 - 1	926-K1-WS-073124	6	-0.0571	-	7470A	PPB	13 Aug 2024 10:16:46	U	SMPL
P3429-02 - 1	931-K1-WS-073124	309	0.0171	-	7470A	PPB	13 Aug 2024 10:19:02	U	SMPL
P3429-03 - 1	925-K1-WS-073124	778	0.1318	-	7470A	PPB	13 Aug 2024 10:21:19	U	SMPL
P3440-01 - 1	923-K1-WS-080124	-109	-0.0852	-	7470A	PPB	13 Aug 2024 10:23:35	U	SMPL
P3440-01DUP - 1	923-K1-WS-080124DUP	17	-0.0544	-	7470A	PPB	13 Aug 2024 10:25:53	U	SMPL
P3440-02 - 1	P3440-01MS	13643	3.2798	-	7470A	PPB	13 Aug 2024 10:28:09	U	SMPL
CCV38 - 1	CCV38	20089	4.8571	-	7470A	PPB	13 Aug 2024 10:30:24	U	SMPL
CCB38 - 1	CCB38	14	-0.0551	-	7470A	PPB	13 Aug 2024 10:35:27	U	SMPL
P3440-03 - 1	P3440-01MSD	13205	3.1727	-	7470A	PPB	13 Aug 2024 10:37:42	U	SMPL
P3440-04 - 1	922-K1-WS-080124	-73	-0.0764	-	7470A	PPB	13 Aug 2024 10:39:57	U	SMPL
P3451-01 - 1	921-J-WS-080124	37	-0.0495	-	7470A	PPB	13 Aug 2024 10:42:15	U	SMPL
P3457-01 - 1	924-K1-WS-080224	100	-0.0341	-	7470A	PPB	13 Aug 2024 10:44:30	U	SMPL
P3457-02 - 1	932-K1-WS-080224	141	-0.0240	-	7470A	PPB	13 Aug 2024 10:46:46	U	SMPL
P3467-01 - 1	919-J-WS-080224	1257	0.2490	-	7470A	PPB	13 Aug 2024 10:49:02	U	SMPL
P3497-02 - 1	COMP	33	-0.0505	-	7470A	PPB	13 Aug 2024 10:51:19	U	SMPL
P3532-01 - 1	806-C	950	0.1739	-	7470A	PPB	13 Aug 2024 10:53:36	U	SMPL
P3545-01 - 1	BUR-1295	914	0.1651	-	7470A	PPB	13 Aug 2024 10:55:52	U	SMPL
CCV39 - 1	CCV39	21395	5.1767	-	7470A	PPB	13 Aug 2024 10:58:08	U	SMPL
CCB39 - 1	CCB39	221	-0.0045	-	7470A	PPB	13 Aug 2024 11:04:09	U	SMPL
PB162588TB - 1	PB162588TB	-23	-0.0642	-	7470A	PPB	13 Aug 2024 11:06:24	U	SMPL
PB162590TB - 1	PB162590TB	-2	-0.0590	-	7470A	PPB	13 Aug 2024 11:08:39	U	SMPL
P3534-03LX5 - 1		-32	-0.0664	-	7470A	PPB	13 Aug 2024 11:10:54	U	SMPL
P3534-03A - 1		15658	3.7729	-	7470A	PPB	13 Aug 2024 11:13:10	U	SMPL
P3530-02LX5 - 1		-459	-0.1709	-	7470A	PPB	13 Aug 2024 11:15:25	U	SMPL
P3530-02A - 1		16246	3.9168	-	7470A	PPB	13 Aug 2024 11:17:42	U	SMPL
P3440-01LX5 - 1		-388	-0.1535	-	7470A	PPB	13 Aug 2024 11:19:58	U	SMPL
P3440-01A - 1		16290	3.9275	-	7470A	PPB	13 Aug 2024 11:22:17	U	SMPL
CCV40 - 1	CCV40	21055	5.0935	-	7470A	PPB	13 Aug 2024 11:24:33	U	SMPL
CCB40 - 1	CCB40	-469	-0.1733	-	7470A	PPB	13 Aug 2024 11:26:51	U	SMPL