

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3429
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3429 **SAS No.:** P3429
Matrix: Water **Level:** LOW **Client ID:** 923-K1-WS-080124A
Sample ID: P3440-01 **Spiked ID:** P3440-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Strontium	ug/L	90 - 113	210		129		500	16		P