

Instrument ID: BNA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # BN071124**

|              |          |                   |                                           |
|--------------|----------|-------------------|-------------------------------------------|
| Review By    | yogesh   | Review On         | 7/11/2024 4:20:04 AM                      |
| Supervise By | mohammad | Supervise On      | 7/11/2024 5:19:17 AM                      |
| SubDirectory | BN071124 | HP Acquire Method | BNA_N, 8270_HP Processing Method bn071124 |

| STD. NAME                | STD REF.#                                        |
|--------------------------|--------------------------------------------------|
| Tune/Reschk              | SP6542                                           |
| Initial Calibration Stds | SP6569,SP6568,SP6567,SP6566,SP6565,SP6564,SP6563 |
| CCC                      | SP6567                                           |
| Internal Standard/PEM    | SP6441                                           |
| ICV/I.BLK                | SP6548                                           |
| Surrogate Standard       |                                                  |
| MS/MSD Standard          |                                                  |
| LCS Standard             |                                                  |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                                   | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|-------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BN032615.D     | 10 Jul 2024 15:03 |                                           | MA/JU    | Ok     |
| 2   | SSTDICC0.1  | SSTDICC0.1  | BN032616.D     | 10 Jul 2024 15:54 | Compound #20 removed from 0.1ppm          | MA/JU    | Ok,M   |
| 3   | SSTDICC0.2  | SSTDICC0.2  | BN032617.D     | 10 Jul 2024 16:30 | Compound #20 Kept on LR                   | MA/JU    | Ok,M   |
| 4   | SSTDICCC0.4 | SSTDICCC0.4 | BN032618.D     | 10 Jul 2024 17:07 | Calibration is Good For DOD except Com#20 | MA/JU    | Ok,M   |
| 5   | SSTDICC0.8  | SSTDICC0.8  | BN032619.D     | 10 Jul 2024 17:43 | Com#20 failed for %RE in calibration      | MA/JU    | Ok,M   |
| 6   | SSTDICC1.6  | SSTDICC1.6  | BN032620.D     | 10 Jul 2024 18:19 |                                           | MA/JU    | Ok,M   |
| 7   | SSTDICC3.2  | SSTDICC3.2  | BN032621.D     | 10 Jul 2024 18:56 |                                           | MA/JU    | Ok,M   |
| 8   | SSTDICC5.0  | SSTDICC5.0  | BN032622.D     | 10 Jul 2024 19:32 |                                           | MA/JU    | Ok,M   |
| 9   | SSTDICV0.4  | ICVBN071124 | BN032623.D     | 10 Jul 2024 20:44 |                                           | MA/JU    | Ok,M   |
| 10  | PB161896BL  | PB161896BL  | BN032624.D     | 10 Jul 2024 21:21 | Analyzed for contamination check          | MA/JU    | Not Ok |

M : Manual Integration

Instrument ID: BNA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # BN080324**

|              |          |                   |                                           |
|--------------|----------|-------------------|-------------------------------------------|
| Review By    | yogesh   | Review On         | 8/5/2024 2:04:10 AM                       |
| Supervise By | mohammad | Supervise On      | 8/7/2024 2:40:34 AM                       |
| SubDirectory | BN080324 | HP Acquire Method | BNA_N, 8270_HP Processing Method bn071124 |

| STD. NAME                | STD REF.#                                        |
|--------------------------|--------------------------------------------------|
| Tune/Reschk              | SP6573                                           |
| Initial Calibration Stds | SP6569,SP6568,SP6567,SP6566,SP6565,SP6564,SP6563 |
| CCC                      | SP6567                                           |
| Internal Standard/PEM    | SP6527                                           |
| ICV/I.BLK                | SP6548                                           |
| Surrogate Standard       |                                                  |
| MS/MSD Standard          |                                                  |
| LCS Standard             |                                                  |

| Sr# | SampleID    | ClientID            | Data File Name | Date-Time         | Comment                | Operator | Status   |
|-----|-------------|---------------------|----------------|-------------------|------------------------|----------|----------|
| 1   | DFTPP       | DFTPP               | BN033211.D     | 02 Aug 2024 20:20 |                        | MA/JU    | Ok       |
| 2   | SSTDCCC0.4  | SSTDCCC0.4          | BN033212.D     | 02 Aug 2024 20:59 |                        | MA/JU    | Ok,M     |
| 3   | PB162424BL  | PB162424BL          | BN033213.D     | 02 Aug 2024 22:12 |                        | MA/JU    | Ok,M     |
| 4   | PB162424BS  | PB162424BS          | BN033214.D     | 02 Aug 2024 22:48 |                        | MA/JU    | Ok,M     |
| 5   | P3410-01    | GW-S-864-9-14-07302 | BN033215.D     | 02 Aug 2024 23:25 | Need 5X Dilution       | MA/JU    | Dilution |
| 6   | P3415-03    | MLS-15-70-85        | BN033216.D     | 03 Aug 2024 00:02 |                        | MA/JU    | Ok,M     |
| 7   | P3415-04MS  | MLS-15-70-85MS      | BN033217.D     | 03 Aug 2024 00:38 |                        | MA/JU    | Ok,M     |
| 8   | P3415-05MSD | MLS-15-70-85MSD     | BN033218.D     | 03 Aug 2024 01:15 |                        | MA/JU    | Ok,M     |
| 9   | P3416-01    | MLS-15-106-121      | BN033219.D     | 03 Aug 2024 01:51 |                        | MA/JU    | Ok,M     |
| 10  | P3416-02    | MLS-15-135-150      | BN033220.D     | 03 Aug 2024 02:27 |                        | MA/JU    | Ok,M     |
| 11  | P3416-03    | MLS-15-9999         | BN033221.D     | 03 Aug 2024 03:03 |                        | MA/JU    | Ok,M     |
| 12  | P3426-01    | 927-K1-WS-073124    | BN033222.D     | 03 Aug 2024 03:39 |                        | MA/JU    | Ok,M     |
| 13  | P3426-02    | 927-K1-WS-073124-FD | BN033223.D     | 03 Aug 2024 04:15 | Internal Standard Fail | MA/JU    | ReRun    |
| 14  | P3429-01    | 926-K1-WS-073124    | BN033224.D     | 03 Aug 2024 04:52 |                        | MA/JU    | Ok,M     |
| 15  | P3429-02    | 931-K1-WS-073124    | BN033225.D     | 03 Aug 2024 05:28 |                        | MA/JU    | Ok,M     |
| 16  | P3429-03    | 925-K1-WS-073124    | BN033226.D     | 03 Aug 2024 06:04 |                        | MA/JU    | Ok,M     |
| 17  | DFTPP       | DFTPP               | BN033227.D     | 03 Aug 2024 07:20 |                        | MA/JU    | Ok       |
| 18  | SSTDCCC0.4  | SSTDCCC0.4          | BN033228.D     | 03 Aug 2024 07:59 |                        | MA/JU    | Ok,M     |

Instrument ID: BNA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # BN080324**

|              |          |                   |                                           |
|--------------|----------|-------------------|-------------------------------------------|
| Review By    | yogesh   | Review On         | 8/5/2024 2:04:10 AM                       |
| Supervise By | mohammad | Supervise On      | 8/7/2024 2:40:34 AM                       |
| SubDirectory | BN080324 | HP Acquire Method | BNA_N, 8270_HP Processing Method bn071124 |

| STD. NAME                | STD REF.#                                        |
|--------------------------|--------------------------------------------------|
| Tune/Reschk              | SP6573                                           |
| Initial Calibration Stds | SP6569,SP6568,SP6567,SP6566,SP6565,SP6564,SP6563 |
| CCC                      | SP6567                                           |
| Internal Standard/PEM    | SP6527                                           |
| ICV/I.BLK                | SP6548                                           |
| Surrogate Standard       |                                                  |
| MS/MSD Standard          |                                                  |
| LCS Standard             |                                                  |

|    |             |                      |            |                   |                            |       |          |
|----|-------------|----------------------|------------|-------------------|----------------------------|-------|----------|
| 19 | PB162464BL  | PB162464BL           | BN033229.D | 03 Aug 2024 09:11 | END CCC Missing            | MA/JU | Not Ok   |
| 20 | PB162464BS  | PB162464BS           | BN033230.D | 03 Aug 2024 09:47 | END CCC Missing            | MA/JU | Not Ok   |
| 21 | P3440-01    | 923-K1-WS-080124     | BN033231.D | 03 Aug 2024 10:24 |                            | MA/JU | Ok,M     |
| 22 | P3440-02MS  | 923-K1-WS-080124MS   | BN033232.D | 03 Aug 2024 11:00 | Used for NON DOD           | MA/JU | Ok,M     |
| 23 | P3440-03MSD | 923-K1-WS-080124MS   | BN033233.D | 03 Aug 2024 11:36 | Used for NON DOD           | MA/JU | Ok,M     |
| 24 | P3440-04    | 922-K1-WS-080124     | BN033234.D | 03 Aug 2024 12:12 |                            | MA/JU | Ok,M     |
| 25 | P3451-01    | 921-J-WS-080124      | BN033235.D | 03 Aug 2024 12:48 |                            | MA/JU | Ok,M     |
| 26 | P3410-01DL  | GW-S-864-9-14-073024 | BN033236.D | 03 Aug 2024 13:24 |                            | MA/JU | Ok,M     |
| 27 | P3426-02RE  | 927-K1-WS-073124-FD  | BN033237.D | 03 Aug 2024 14:00 | Internal Standard Fail     | MA/JU | Confirms |
| 28 | SP6586      | SP6586               | BN033238.D | 03 Aug 2024 14:36 | SP6586--8270-SIM SURROGATE | MA/JU | Ok,M     |

M : Manual Integration