

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071124\
 Data File : BN032622.D
 Acq On : 10 Jul 2024 19:32
 Operator : MA/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Jul 10 23:14:56 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 10 23:11:23 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.559	152	5387	0.400	ng	0.00
7) Naphthalene-d8	10.325	136	16995	0.400	ng	-0.03
13) Acenaphthene-d10	14.199	164	11270	0.400	ng	-0.01
19) Phenanthrene-d10	16.967	188	23755	0.400	ng	#-0.01
29) Chrysene-d12	21.175	240	21678	0.400	ng	#-0.02
35) Perylene-d12	23.382	264	19827	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.176	112	63938	4.697	ng	-0.02
5) Phenol-d6	6.758	99	88108	5.063	ng	-0.04
8) Nitrobenzene-d5	8.723	82	65619	5.172	ng	-0.05
11) 2-Methylnaphthalene-d10	11.926	152	130262	4.850	ng	-0.04
14) 2,4,6-Tribromophenol	15.713	330	27571	5.399	ng	-0.02
15) 2-Fluorobiphenyl	12.820	172	204044	4.914	ng	-0.03
27) Fluoranthene-d10	19.003	212	310962	5.023	ng	-0.02
31) Terphenyl-d14	19.611	244	207960	3.929	ng	-0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.132	88	27696	3.904	ng	98
3) n-Nitrosodimethylamine	3.450	42	25796	5.010	ng	99
6) bis(2-Chloroethyl)ether	7.003	93	74029	6.194	ng	90
9) Naphthalene	10.378	128	227198	4.894	ng	96
10) Hexachlorobutadiene	10.645	225	41068	4.441	ng	# 99
12) 2-Methylnaphthalene	12.002	142	156743	5.044	ng	97
16) Acenaphthylene	13.921	152	246134	5.072	ng	99
17) Acenaphthene	14.264	154	158672	4.704	ng	99
18) Fluorene	15.258	166	212366	4.805	ng	100
21) 4-Bromophenyl-phenylether	16.160	248	69439	4.922	ng	# 74
22) Hexachlorobenzene	16.259	284	73899	4.762	ng	100
23) Atrazine	16.445	200	56576	5.136	ng	# 93
24) Pentachlorophenol	16.631	266	34124	6.557	ng	98
25) Phenanthrene	17.004	178	325236	5.040	ng	100
26) Anthracene	17.091	178	302632	5.831	ng	99
28) Fluoranthene	19.035	202	399794	5.128	ng	100
30) Pyrene	19.402	202	407347	4.619	ng	99
32) Benzo(a)anthracene	21.157	228	366789	5.115	ng	100
33) Chrysene	21.211	228	382139	4.488	ng	99
34) Bis(2-ethylhexyl)phtha...	21.077	149	217511	4.464	ng	99
36) Indeno(1,2,3-cd)pyrene	25.580	276	381332	5.034	ng	99
37) Benzo(b)fluoranthene	22.712	252	374321	5.665	ng	# 86
38) Benzo(k)fluoranthene	22.756	252	375339	4.931	ng	# 85
39) Benzo(a)pyrene	23.282	252	315369	5.042	ng	# 77
40) Dibenzo(a,h)anthracene	25.586	278	304385	5.063	ng	# 90
41) Benzo(g,h,i)perylene	26.262	276	320415	4.895	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071124\
Data File : BN032622.D
Acq On : 10 Jul 2024 19:32
Operator : MA/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jul 10 23:14:56 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 10 23:11:23 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.559 min Scan# 61
 Delta R.T. -0.008 min
 Lab File: BN032622.D
 Acq: 10 Jul 2024 19:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:152 Resp: 5387
 Ion Ratio Lower Upper
 152 100
 150 153.7 123.8 185.6
 115 60.8 52.8 79.2



#2
 1,4-Dioxane
 Concen: 3.904 ng
 RT: 3.132 min Scan# 21
 Delta R.T. -0.007 min
 Lab File: BN032622.D
 Acq: 10 Jul 2024 19:32

Tgt Ion: 88 Resp: 27696
 Ion Ratio Lower Upper
 88 100
 43 22.7 18.4 27.6
 58 63.8 49.2 73.8









#3
n-Nitrosodimethylamine
Concen: 5.010 ng
RT: 3.450 min Scan# 61
Delta R.T. -0.022 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 25796
Ion Ratio Lower Upper
42 100
74 159.7 126.9 190.3
44 6.4 4.8 7.2



#4
2-Fluorophenol
Concen: 4.697 ng
RT: 5.176 min Scan# 304
Delta R.T. -0.022 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion: 112 Resp: 63938
Ion Ratio Lower Upper
112 100
64 51.2 40.2 60.4
63 26.3 22.2 33.4









#5
Phenol-d6
Concen: 5.063 ng
RT: 6.758 min Scan# 51
Delta R.T. -0.036 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 88108
Ion Ratio Lower Upper
99 100
42 14.9 10.5 15.7
71 31.2 22.3 33.5



#6
bis(2-Chloroethyl)ether
Concen: 6.194 ng
RT: 7.003 min Scan# 557
Delta R.T. -0.014 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion: 93 Resp: 74029
Ion Ratio Lower Upper
93 100
63 68.6 46.4 69.6
95 33.3 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.325 min Scan# 91
Delta R.T. -0.032 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	16995
Ion	Ratio	Lower	Upper
136	100		
137	11.7	9.9	14.9
54	7.9	6.5	9.7
68	7.2	5.8	8.8



#8
Nitrobenzene-d5
Concen: 5.172 ng
RT: 8.723 min Scan# 777
Delta R.T. -0.053 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:	82	Resp:	65619
Ion	Ratio	Lower	Upper
82	100		
128	48.6	41.7	62.5
54	39.7	39.5	59.3









#9
Naphthalene
Concen: 4.894 ng
RT: 10.378 min Scan# 91
Delta R.T. -0.021 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 227198
Ion Ratio Lower Upper
128 100
129 10.9 9.8 14.8
127 12.9 11.5 17.3



#10
Hexachlorobutadiene
Concen: 4.441 ng
RT: 10.645 min Scan# 957
Delta R.T. 0.000 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:225 Resp: 41068
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.3 76.9









#11
2-Methylnaphthalene-d10
Concen: 4.850 ng
RT: 11.926 min Scan# 1158
Delta R.T. -0.038 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 130262
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2



#12
2-Methylnaphthalene
Concen: 5.044 ng
RT: 12.002 min Scan# 1168
Delta R.T. -0.038 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:142 Resp: 156743
Ion Ratio Lower Upper
142 100
141 89.3 69.3 103.9
115 35.2 29.5 44.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.199 min Scan# 14
Delta R.T. -0.011 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 11270
Ion Ratio Lower Upper
164 100
162 105.5 85.0 127.4
160 49.1 39.4 59.2



#14
2,4,6-Tribromophenol
Concen: 5.399 ng
RT: 15.713 min Scan# 1616
Delta R.T. -0.025 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:330 Resp: 27571
Ion Ratio Lower Upper
330 100
332 96.1 77.8 116.8
141 39.0 34.2 51.2





#15

2-Fluorobiphenyl

Concen: 4.914 ng

RT: 12.820 min Scan# 1348

Delta R.T. -0.032 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:172 Resp: 204044

Ion Ratio Lower Upper

172 100

171 34.4 28.6 43.0

170 22.3 19.0 28.6



#16

Acenaphthylene

Concen: 5.072 ng

RT: 13.921 min Scan# 1451

Delta R.T. -0.011 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32




Tgt Ion:152 Resp: 246134


Ion Ratio Lower Upper

152 100

151 19.4 16.0 24.0

153 13.0 10.6 15.8





#17

Acenaphthene

Concen: 4.704 ng

RT: 14.264 min Scan# 1483

Delta R.T. -0.011 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:154 Resp: 158672

Ion Ratio Lower Upper

154 100

153 114.1 91.8 137.8

152 55.4 45.2 67.8



#18

Fluorene

Concen: 4.805 ng

RT: 15.258 min Scan# 1576

Delta R.T. -0.021 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32




Tgt Ion:166 Resp: 212366


Ion Ratio Lower Upper

166 100

165 97.5 78.0 117.0

167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.967 min Scan# 1718
Delta R.T. -0.012 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 23755
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 8.8 13.2#



#21
4-Bromophenyl-phenylether
Concen: 4.922 ng
RT: 16.160 min Scan# 1652
Delta R.T. -0.012 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:248 Resp: 69439
Ion Ratio Lower Upper
248 100
250 102.8 75.1 112.7
141 37.0 62.0 93.0#









#22
Hexachlorobenzene
Concen: 4.762 ng
RT: 16.259 min Scan# 1660
Delta R.T. -0.012 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 73899
Ion Ratio Lower Upper
284 100
142 36.5 29.4 44.0
249 30.6 24.6 36.8



#23
Atrazine
Concen: 5.136 ng
RT: 16.445 min Scan# 1675
Delta R.T. -0.012 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:200 Resp: 56576
Ion Ratio Lower Upper
200 100
173 23.2 25.4 38.2#
215 48.5 38.2 57.2









#24
Pentachlorophenol
Concen: 6.557 ng
RT: 16.631 min Scan# 1690
Delta R.T. -0.025 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:266 Resp: 34124
Ion Ratio Lower Upper
266 100
264 62.8 50.2 75.2
268 63.7 52.8 79.2



#25
Phenanthrene
Concen: 5.040 ng
RT: 17.004 min Scan# 1720
Delta R.T. -0.012 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:178 Resp: 325236
Ion Ratio Lower Upper
178 100
176 19.1 15.1 22.7
179 15.2 12.2 18.4





#26

Anthracene

Concen: 5.831 ng

RT: 17.091 min Scan# 1

Delta R.T. -0.037 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:178 Resp: 302632

Ion Ratio Lower Upper

178 100

176 18.4 14.8 22.2

179 15.2 11.8 17.8



#27

Fluoranthene-d10

Concen: 5.023 ng

RT: 19.003 min Scan# 1948

Delta R.T. -0.019 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32




Tgt Ion:212 Resp: 310962


Ion Ratio Lower Upper

212 100

106 14.9 11.9 17.9

104 8.5 6.9 10.3





#28

Fluoranthene

Concen: 5.128 ng

RT: 19.035 min Scan# 1955

Delta R.T. -0.019 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:202 Resp: 399794

Ion Ratio Lower Upper

202 100

101 12.0 9.6 14.4

203 17.2 13.8 20.6



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.175 min Scan# 2398

Delta R.T. -0.018 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32




Tgt Ion:240 Resp: 21678


Ion Ratio Lower Upper

240 100

120 12.8 14.8 22.2#

236 28.1 23.3 34.9









#30
Pyrene
Concen: 4.619 ng
RT: 19.402 min Scan# 2034
Delta R.T. -0.019 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 407347
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.9 14.8 22.2



#31
Terphenyl-d14
Concen: 3.929 ng
RT: 19.611 min Scan# 2079
Delta R.T. -0.019 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:244 Resp: 207960
Ion Ratio Lower Upper
244 100
212 8.2 7.5 11.3
122 14.4 12.6 19.0









#32
Benzo(a)anthracene
Concen: 5.115 ng
RT: 21.157 min Scan# 21
Delta R.T. -0.027 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 366789
Ion Ratio Lower Upper
228 100
226 27.1 21.8 32.6
229 19.7 16.0 24.0



#33
Chrysene
Concen: 4.488 ng
RT: 21.211 min Scan# 2402
Delta R.T. -0.027 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:228 Resp: 382139
Ion Ratio Lower Upper
228 100
226 29.9 24.2 36.2
229 19.4 16.0 24.0









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 4.464 ng
 RT: 21.077 min Scan# 2107
 Delta R.T. -0.009 min
 Lab File: BN032622.D
 Acq: 10 Jul 2024 19:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:149 Resp: 217511
 Ion Ratio Lower Upper
 149 100
 167 27.6 21.4 32.2
 279 4.0 2.8 4.2



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.382 min Scan# 2868
 Delta R.T. -0.020 min
 Lab File: BN032622.D
 Acq: 10 Jul 2024 19:32

Tgt Ion:264 Resp: 19827
 Ion Ratio Lower Upper
 264 100
 260 25.7 20.9 31.3
 265 76.3 85.4 128.2#





#36

Indeno(1,2,3-cd)pyrene

Concen: 5.034 ng

RT: 25.580 min Scan# 3

Delta R.T. -0.050 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:276 Resp: 381332

Ion Ratio Lower Upper

276 100

138 26.8 21.0 31.4

277 24.9 19.5 29.3



#37

Benzo(b)fluoranthene

Concen: 5.665 ng

RT: 22.712 min Scan# 2639

Delta R.T. -0.026 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32




Tgt Ion:252 Resp: 374321


Ion Ratio Lower Upper

252 100

253 22.0 22.3 33.5#

125 11.1 15.4 23.2#





#38

Benzo(k)fluoranthene

Concen: 4.931 ng

RT: 22.756 min Scan# 2

Delta R.T. -0.026 min

Lab File: BN032622.D


Acq: 10 Jul 2024 19:32

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 375339

Ion Ratio Lower Upper

252 100

253 22.1 22.6 34.0#

125 11.0 16.2 24.4#



#39

Benzo(a)pyrene

Concen: 5.042 ng

RT: 23.282 min Scan# 2834

Delta R.T. -0.032 min

Lab File: BN032622.D

Acq: 10 Jul 2024 19:32




Tgt Ion:252 Resp: 315369


Ion Ratio Lower Upper

252 100

253 22.0 25.8 38.8#

125 12.4 21.4 32.2#









#40
Dibenzo(a,h)anthracene
Concen: 5.063 ng
RT: 25.586 min Scan# 30
Delta R.T. -0.053 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 304385
Ion Ratio Lower Upper
278 100
139 17.9 16.7 25.1
279 23.5 24.3 36.5#



#41
Benzo(g,h,i)perylene
Concen: 4.895 ng
RT: 26.262 min Scan# 3853
Delta R.T. -0.047 min
Lab File: BN032622.D
Acq: 10 Jul 2024 19:32

Tgt Ion:276 Resp: 320415
Ion Ratio Lower Upper
276 100
277 23.7 20.2 30.2
138 23.2 20.5 30.7

