



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: P3429

Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID : 926-K1-WS-073124								
P3429-01	926-K1-WS-073124	WATER	1,4-Dioxane	0.170	J	0.07	0.2	ug/L
Total Svoc :				0.17				
Total Concentration:				0.17				
Client ID : 925-K1-WS-073124								
P3429-03	925-K1-WS-073124	WATER	Fluoranthene	0.030	J	0.02	0.1	ug/L
P3429-03	925-K1-WS-073124	WATER	Pyrene	0.030	J	0.02	0.1	ug/L
P3429-03	925-K1-WS-073124	WATER	Benzo(b)fluoranthene	0.030	J	0.03	0.1	ug/L
P3429-03	925-K1-WS-073124	WATER	1,4-Dioxane	0.440		0.07	0.2	ug/L
Total Svoc :				0.53				
Total Concentration:				0.53				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	07/31/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	07/31/24
Client Sample ID:	926-K1-WS-073124	SDG No.:	P3429
Lab Sample ID:	P3429-01	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN033224.D	1	08/01/24 08:58	08/03/24 04:52	PB162424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	0.020	U	0.020	0.10	ug/L
91-57-6	2-Methylnaphthalene	0.030	U	0.030	0.10	ug/L
208-96-8	Acenaphthylene	0.020	U	0.020	0.10	ug/L
83-32-9	Acenaphthene	0.020	U	0.020	0.10	ug/L
86-73-7	Fluorene	0.020	U	0.020	0.10	ug/L
85-01-8	Phenanthrene	0.020	U	0.020	0.10	ug/L
120-12-7	Anthracene	0.020	U	0.020	0.10	ug/L
206-44-0	Fluoranthene	0.020	U	0.020	0.10	ug/L
129-00-0	Pyrene	0.020	U	0.020	0.10	ug/L
56-55-3	Benzo(a)anthracene	0.020	U	0.020	0.10	ug/L
218-01-9	Chrysene	0.030	U	0.030	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.030	U	0.030	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.030	U	0.030	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.060	U	0.060	0.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	0.040	U	0.040	0.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	0.040	U	0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.040	U	0.040	0.10	ug/L
123-91-1	1,4-Dioxane	0.17	J	0.070	0.20	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.27		30 (30) - 150 (150)	67%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.29		30 (30) - 150 (150)	72%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		30 (11) - 130 (175)	86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.30		30 (10) - 130 (175)	75%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		30 (54) - 130 (171)	91%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2820	7.546			
1146-65-2	Naphthalene-d8	9530	10.308			
15067-26-2	Acenaphthene-d10	5490	14.143			
1517-22-2	Phenanthrene-d10	10800	16.914			

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	07/31/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	07/31/24
Client Sample ID:	926-K1-WS-073124	SDG No.:	P3429
Lab Sample ID:	P3429-01	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN033224.D	1	08/01/24 08:58	08/03/24 04:52	PB162424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
1719-03-5	Chrysene-d12	8090	21.131			
1520-96-3	Perylene-d12	8930	23.301			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	07/31/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	07/31/24
Client Sample ID:	931-K1-WS-073124	SDG No.:	P3429
Lab Sample ID:	P3429-02	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN033225.D	1	08/01/24 08:58	08/03/24 05:28	PB162424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	0.020	U	0.020	0.10	ug/L
91-57-6	2-Methylnaphthalene	0.030	U	0.030	0.10	ug/L
208-96-8	Acenaphthylene	0.020	U	0.020	0.10	ug/L
83-32-9	Acenaphthene	0.020	U	0.020	0.10	ug/L
86-73-7	Fluorene	0.020	U	0.020	0.10	ug/L
85-01-8	Phenanthrene	0.020	U	0.020	0.10	ug/L
120-12-7	Anthracene	0.020	U	0.020	0.10	ug/L
206-44-0	Fluoranthene	0.020	U	0.020	0.10	ug/L
129-00-0	Pyrene	0.020	U	0.020	0.10	ug/L
56-55-3	Benzo(a)anthracene	0.020	U	0.020	0.10	ug/L
218-01-9	Chrysene	0.030	U	0.030	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.030	U	0.030	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.030	U	0.030	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.060	U	0.060	0.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	0.040	U	0.040	0.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	0.040	U	0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.040	U	0.040	0.10	ug/L
123-91-1	1,4-Dioxane	0.070	U	0.070	0.20	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.27		30 (30) - 150 (150)	68%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.31		30 (30) - 150 (150)	76%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		30 (11) - 130 (175)	90%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		30 (10) - 130 (175)	84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		30 (54) - 130 (171)	97%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2740	7.517			
1146-65-2	Naphthalene-d8	9930	10.265			
15067-26-2	Acenaphthene-d10	5090	14.137			
1517-22-2	Phenanthrene-d10	10200	16.908			

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	07/31/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	07/31/24
Client Sample ID:	931-K1-WS-073124	SDG No.:	P3429
Lab Sample ID:	P3429-02	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN033225.D	1	08/01/24 08:58	08/03/24 05:28	PB162424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
1719-03-5	Chrysene-d12	8230	21.125			
1520-96-3	Perylene-d12	9150	23.303			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	07/31/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	07/31/24
Client Sample ID:	925-K1-WS-073124	SDG No.:	P3429
Lab Sample ID:	P3429-03	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN033226.D	1	08/01/24 08:58	08/03/24 06:04	PB162424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	0.020	U	0.020	0.10	ug/L
91-57-6	2-Methylnaphthalene	0.030	U	0.030	0.10	ug/L
208-96-8	Acenaphthylene	0.020	U	0.020	0.10	ug/L
83-32-9	Acenaphthene	0.020	U	0.020	0.10	ug/L
86-73-7	Fluorene	0.020	U	0.020	0.10	ug/L
85-01-8	Phenanthrene	0.020	U	0.020	0.10	ug/L
120-12-7	Anthracene	0.020	U	0.020	0.10	ug/L
206-44-0	Fluoranthene	0.030	J	0.020	0.10	ug/L
129-00-0	Pyrene	0.030	J	0.020	0.10	ug/L
56-55-3	Benzo(a)anthracene	0.020	U	0.020	0.10	ug/L
218-01-9	Chrysene	0.030	U	0.030	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.030	J	0.030	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.030	U	0.030	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.060	U	0.060	0.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	0.040	U	0.040	0.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	0.040	U	0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.040	U	0.040	0.10	ug/L
123-91-1	1,4-Dioxane	0.44		0.070	0.20	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.29		30 (30) - 150 (150)	73%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 (30) - 150 (150)	81%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.22		30 (11) - 130 (175)	54%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		30 (10) - 130 (175)	84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.44		30 (54) - 130 (171)	111%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2640	7.532			
1146-65-2	Naphthalene-d8	9710	10.276			
15067-26-2	Acenaphthene-d10	5270	14.137			
1517-22-2	Phenanthrene-d10	11000	16.908			

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	07/31/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	07/31/24
Client Sample ID:	925-K1-WS-073124	SDG No.:	P3429
Lab Sample ID:	P3429-03	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN033226.D	1	08/01/24 08:58	08/03/24 06:04	PB162424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
1719-03-5	Chrysene-d12	7650	21.125			
1520-96-3	Perylene-d12	8320	23.303			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	P3429	OrderDate:	7/31/2024 4:38:00 PM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger Site Princeton NJ
Contact:	Mary I. Murphy	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P3429-01	926-K1-WS-073124	Water	SVOC-SIMGroup1	8270-Modified	07/31/24	08/01/24	08/03/24	07/31/24
P3429-02	931-K1-WS-073124	Water	SVOC-SIMGroup1	8270-Modified	07/31/24	08/01/24	08/03/24	07/31/24
P3429-03	925-K1-WS-073124	Water	SVOC-SIMGroup1	8270-Modified	07/31/24	08/01/24	08/03/24	07/31/24