

Surrogate Summary

SW-846

SDG No.: P3429

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3415-04MS	MLS-15-70-85MS	2-Methylnaphthalene-d10	0.4	0.34	84		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.37	91		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.37	93		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.36	90		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.52	129		30 (54)	130 (171)
P3415-05MSD	MLS-15-70-85MSD	2-Methylnaphthalene-d10	0.4	0.31	78		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.33	82		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.34	85		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.34	84		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.49	122		30 (54)	130 (171)
P3429-01	926-K1-WS-073124	2-Methylnaphthalene-d10	0.4	0.27	67		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.29	72		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.35	86		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.30	75		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.36	91		30 (54)	130 (171)
P3429-02	931-K1-WS-073124	2-Methylnaphthalene-d10	0.4	0.27	68		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.31	76		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.36	90		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.34	84		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.39	97		30 (54)	130 (171)
P3429-03	925-K1-WS-073124	2-Methylnaphthalene-d10	0.4	0.29	73		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.33	81		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.22	54		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.34	84		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.44	111		30 (54)	130 (171)
PB162424BL	PB162424BL	2-Methylnaphthalene-d10	0.4	0.31	78		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.32	81		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.38	95		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.34	84		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.37	92		30 (54)	130 (171)
PB162424BS	PB162424BS	2-Methylnaphthalene-d10	0.4	0.35	88		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.33	81		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.39	97		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.35	87		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.35	88		30 (54)	130 (171)