

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF073024**

|              |          |                      |                      |
|--------------|----------|----------------------|----------------------|
| Review By    | yogesh   | Review On            | 7/31/2024 6:39:45 AM |
| Supervise By | mohammad | Supervise On         | 7/31/2024 8:08:34 AM |
| SubDirectory | BF073024 | HP Acquire Method    | BNA_F                |
|              |          | HP Processing Method | bf073024             |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6573                                                  |
| Initial Calibration Stds | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |
| CCC                      | SP6553                                                  |
| Internal Standard/PEM    | S12036 10ul/1000ul sample                               |
| ICV/I.BLK                | SP6559                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                                       | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|-----------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BF138679.D     | 30 Jul 2024 12:24 |                                               | RC/JU    | Ok     |
| 2   | SSTDICC2.5  | SSTDICC2.5  | BF138680.D     | 30 Jul 2024 12:54 |                                               | RC/JU    | Ok     |
| 3   | SSTDICC005  | SSTDICC005  | BF138681.D     | 30 Jul 2024 13:25 | Compound#32,41,54,56,65,70 removed from 5 ppm | RC/JU    | Ok     |
| 4   | SSTDICC010  | SSTDICC010  | BF138682.D     | 30 Jul 2024 13:56 |                                               | RC/JU    | Ok,M   |
| 5   | SSTDICC020  | SSTDICC020  | BF138683.D     | 30 Jul 2024 14:25 | Comopund#41 Kept on LR                        | RC/JU    | Ok     |
| 6   | SSTDICCC040 | SSTDICCC040 | BF138684.D     | 30 Jul 2024 14:56 | The calibration is good for 8270 DOD & 625.1  | RC/JU    | Ok     |
| 7   | SSTDICC050  | SSTDICC050  | BF138685.D     | 30 Jul 2024 15:27 |                                               | RC/JU    | Ok     |
| 8   | SSTDICC060  | SSTDICC060  | BF138686.D     | 30 Jul 2024 15:58 | Compound#9 removed from 60 ppm                | RC/JU    | Ok     |
| 9   | SSTDICC080  | SSTDICC080  | BF138687.D     | 30 Jul 2024 16:29 | Compound#9,77 removed from 80 ppm             | RC/JU    | Ok     |
| 10  | SSTDICV040  | ICVBF073024 | BF138688.D     | 30 Jul 2024 17:55 |                                               | RC/JU    | Ok     |
| 11  | PB162188BL  | PB162188BL  | BF138689.D     | 30 Jul 2024 18:26 |                                               | RC/JU    | Ok     |

M : Manual Integration

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF080724**

|              |          |                      |                     |
|--------------|----------|----------------------|---------------------|
| Review By    | yogesh   | Review On            | 8/8/2024 8:44:16 AM |
| Supervise By | mohammad | Supervise On         | 8/8/2024 9:27:02 AM |
| SubDirectory | BF080724 | HP Acquire Method    | BNA_F               |
|              |          | HP Processing Method | bf073024            |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6573                                                  |
| Initial Calibration Stds | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |
| CCC                      | SP6553                                                  |
| Internal Standard/PEM    | S12038 10ul/1000ul sample                               |
| ICV/I.BLK                | SP6559                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID                 | Data File Name | Date-Time         | Comment        | Operator | Status |
|-----|-------------|--------------------------|----------------|-------------------|----------------|----------|--------|
| 1   | DFTPP       | DFTPP                    | BF138833.D     | 07 Aug 2024 10:30 |                | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040               | BF138834.D     | 07 Aug 2024 11:00 |                | RC/JU    | Ok     |
| 3   | PB162358BL  | PB162358BL               | BF138835.D     | 07 Aug 2024 11:29 |                | RC/JU    | Ok     |
| 4   | PB162463BL  | PB162463BL               | BF138836.D     | 07 Aug 2024 11:59 |                | RC/JU    | Ok     |
| 5   | P3440-01    | 923-K1-WS-080124         | BF138837.D     | 07 Aug 2024 12:34 |                | RC/JU    | Ok     |
| 6   | P3440-02MS  | 923-K1-WS-080124MS       | BF138838.D     | 07 Aug 2024 13:04 |                | RC/JU    | Ok,M   |
| 7   | P3440-03MSD | 923-K1-WS-080124MS       | BF138839.D     | 07 Aug 2024 13:35 |                | RC/JU    | Ok     |
| 8   | P3440-04    | 922-K1-WS-080124         | BF138840.D     | 07 Aug 2024 14:05 |                | RC/JU    | Ok     |
| 9   | P3450-01    | 921-J-WP0-0.25-080124    | BF138841.D     | 07 Aug 2024 14:34 |                | RC/JU    | Ok     |
| 10  | P3450-02    | 923-K1-WP0-0.25-080124   | BF138842.D     | 07 Aug 2024 15:04 |                | RC/JU    | Ok     |
| 11  | P3450-03    | 922-K1-WP0-0.25-080124   | BF138843.D     | 07 Aug 2024 15:34 |                | RC/JU    | Ok     |
| 12  | P3451-01    | 921-J-WS-080124          | BF138844.D     | 07 Aug 2024 16:05 |                | RC/JU    | Ok     |
| 13  | P3426-01    | 927-K1-WS-073124         | BF138845.D     | 07 Aug 2024 16:35 |                | RC/JU    | Ok     |
| 14  | P3426-02    | 927-K1-WS-073124-FD      | BF138846.D     | 07 Aug 2024 17:05 | Surrogate Fail | RC/JU    | Ok     |
| 15  | P3429-01    | 926-K1-WS-073124         | BF138847.D     | 07 Aug 2024 17:36 |                | RC/JU    | Ok     |
| 16  | P3429-02    | 931-K1-WS-073124         | BF138848.D     | 07 Aug 2024 18:06 |                | RC/JU    | Ok     |
| 17  | P3429-03    | 925-K1-WS-073124         | BF138849.D     | 07 Aug 2024 18:36 |                | RC/JU    | Ok     |
| 18  | P3430-01    | 927-K1-WPO-0-0.25-073124 | BF138850.D     | 07 Aug 2024 19:06 |                | RC/JU    | Ok     |

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF080724**

|              |          |                   |                     |                      |          |
|--------------|----------|-------------------|---------------------|----------------------|----------|
| Review By    | yogesh   | Review On         | 8/8/2024 8:44:16 AM |                      |          |
| Supervise By | mohammad | Supervise On      | 8/8/2024 9:27:02 AM |                      |          |
| SubDirectory | BF080724 | HP Acquire Method | BNA_F               | HP Processing Method | bf073024 |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6573                                                  |
| Initial Calibration Stds | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |
| CCC                      | SP6553                                                  |
| Internal Standard/PEM    | S12038 10ul/1000ul sample                               |
| ICV/I.BLK                | SP6559                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

|    |          |                      |            |                   |  |       |    |
|----|----------|----------------------|------------|-------------------|--|-------|----|
| 19 | P3430-02 | 927-K1-WPO-0-0.25-07 | BF138851.D | 07 Aug 2024 19:37 |  | RC/JU | Ok |
| 20 | P3430-03 | 926-K1-WPO-0-0.25-07 | BF138852.D | 07 Aug 2024 20:07 |  | RC/JU | Ok |
| 21 | P3430-04 | 931-K1-WPO-0-0.25-07 | BF138853.D | 07 Aug 2024 20:37 |  | RC/JU | Ok |
| 22 | P3430-05 | 925-K1-WPO-0-0.25-07 | BF138854.D | 07 Aug 2024 21:08 |  | RC/JU | Ok |

M : Manual Integration

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF080924**

|              |          |                      |                       |
|--------------|----------|----------------------|-----------------------|
| Review By    | yogesh   | Review On            | 8/10/2024 12:55:52 AM |
| Supervise By | mohammad | Supervise On         | 8/12/2024 1:39:31 AM  |
| SubDirectory | BF080924 | HP Acquire Method    | BNA_F                 |
|              |          | HP Processing Method | bf073024              |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6573                                                  |
| Initial Calibration Stds | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |
| CCC                      | SP6553                                                  |
| Internal Standard/PEM    | S12039,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6559                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID       | Data File Name | Date-Time         | Comment                | Operator | Status |
|-----|-------------|----------------|----------------|-------------------|------------------------|----------|--------|
| 1   | DFTPP       | DFTPP          | BF138878.D     | 09 Aug 2024 09:17 |                        | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040     | BF138879.D     | 09 Aug 2024 09:48 |                        | RC/JU    | Ok,M   |
| 3   | PB162421BL  | PB162421BL     | BF138880.D     | 09 Aug 2024 10:19 |                        | RC/JU    | Ok     |
| 4   | PB162358BL  | PB162358BL     | BF138881.D     | 09 Aug 2024 10:50 | Surrogate Fail         | RC/JU    | Not Ok |
| 5   | PB162358BS  | PB162358BS     | BF138882.D     | 09 Aug 2024 11:21 |                        | RC/JU    | Ok,M   |
| 6   | PB162423BL  | PB162423BL     | BF138883.D     | 09 Aug 2024 11:52 |                        | RC/JU    | Ok     |
| 7   | PB162423BS  | PB162423BS     | BF138884.D     | 09 Aug 2024 12:23 |                        | RC/JU    | Ok,M   |
| 8   | PB162489BL  | PB162489BL     | BF138885.D     | 09 Aug 2024 12:55 |                        | RC/JU    | Ok     |
| 9   | PB162489BS  | PB162489BS     | BF138886.D     | 09 Aug 2024 13:25 |                        | RC/JU    | Ok,M   |
| 10  | P3464-02    | TOPSOIL        | BF138887.D     | 09 Aug 2024 14:10 |                        | RC/JU    | Ok     |
| 11  | P3468-01    | IDW-S1-01      | BF138888.D     | 09 Aug 2024 14:59 |                        | RC/JU    | Ok     |
| 12  | P3468-02    | MLS-15-682-697 | BF138889.D     | 09 Aug 2024 15:31 |                        | RC/JU    | Ok     |
| 13  | P3469-01    | MLS-15-513-528 | BF138890.D     | 09 Aug 2024 16:02 |                        | RC/JU    | Ok     |
| 14  | P3469-02    | MLS-15-625-640 | BF138891.D     | 09 Aug 2024 16:33 |                        | RC/JU    | Ok     |
| 15  | P3460-01    | 72-11929       | BF138892.D     | 09 Aug 2024 17:04 |                        | RC/JU    | Ok     |
| 16  | P3460-01MS  | 72-11929MS     | BF138893.D     | 09 Aug 2024 17:35 |                        | RC/JU    | Ok,M   |
| 17  | P3460-01MSD | 72-11929MSD    | BF138894.D     | 09 Aug 2024 18:05 |                        | RC/JU    | Ok,M   |
| 18  | P3507-01    | EO-03-080724   | BF138895.D     | 09 Aug 2024 18:35 | Internal Standard Fail | RC/JU    | Ok,M   |

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF080924**

| Review By                | yogesh                                                  | Review On         | 8/10/2024 12:55:52 AM |                      |          |
|--------------------------|---------------------------------------------------------|-------------------|-----------------------|----------------------|----------|
| Supervise By             | mohammad                                                | Supervise On      | 8/12/2024 1:39:31 AM  |                      |          |
| SubDirectory             | BF080924                                                | HP Acquire Method | BNA_F                 | HP Processing Method | bf073024 |
| STD. NAME                | STD REF.#                                               |                   |                       |                      |          |
| Tune/Reschk              | SP6573                                                  |                   |                       |                      |          |
| Initial Calibration Stds | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |                   |                       |                      |          |
| CCC                      | SP6553                                                  |                   |                       |                      |          |
| Internal Standard/PEM    | S12039,10ul/1000ul sample                               |                   |                       |                      |          |
| ICV/I.BLK                | SP6559                                                  |                   |                       |                      |          |
| Surrogate Standard       |                                                         |                   |                       |                      |          |
| MS/MSD Standard          |                                                         |                   |                       |                      |          |
| LCS Standard             |                                                         |                   |                       |                      |          |

|    |             |                 |            |                   |                        |       |       |
|----|-------------|-----------------|------------|-------------------|------------------------|-------|-------|
| 19 | P3509-01    | OR-02-080724    | BF138896.D | 09 Aug 2024 19:06 | Internal Standard Fail | RC/JU | Ok,M  |
| 20 | P3509-01MS  | OR-02-080724MS  | BF138897.D | 09 Aug 2024 19:36 | Internal Standard Fail | RC/JU | Ok,M  |
| 21 | P3509-01MSD | OR-02-080724MSD | BF138898.D | 09 Aug 2024 20:06 | Internal Standard Fail | RC/JU | Ok,M  |
| 22 | P3394-02    | B-105-SB02      | BF138899.D | 09 Aug 2024 20:36 | Internal Standard Fail | RC/JU | ReRun |

M : Manual Integration

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF081024**

|              |          |                      |                      |
|--------------|----------|----------------------|----------------------|
| Review By    | Jagrut   | Review On            | 8/12/2024 5:10:00 PM |
| Supervise By | mohammad | Supervise On         | 8/13/2024 5:17:21 AM |
| SubDirectory | BF081024 | HP Acquire Method    | BNA_F                |
|              |          | HP Processing Method | bf073024             |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6573                                                  |
| Initial Calibration Stds | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |
| CCC                      | SP6553                                                  |
| Internal Standard/PEM    | S12039,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6559                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID             | Data File Name | Date-Time         | Comment                 | Operator | Status |
|-----|-------------|----------------------|----------------|-------------------|-------------------------|----------|--------|
| 1   | DFTPP       | DFTPP                | BF138900.D     | 10 Aug 2024 10:12 |                         | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040           | BF138901.D     | 10 Aug 2024 10:41 |                         | RC/JU    | Ok,M   |
| 3   | PB162577TB  | PB162577TB           | BF138902.D     | 10 Aug 2024 11:11 |                         | RC/JU    | Ok     |
| 4   | P3479-03    | TS09                 | BF138903.D     | 10 Aug 2024 11:46 |                         | RC/JU    | Ok     |
| 5   | P3479-05    | TS10                 | BF138904.D     | 10 Aug 2024 12:16 |                         | RC/JU    | Ok     |
| 6   | P3479-01    | TS08                 | BF138905.D     | 10 Aug 2024 12:46 |                         | RC/JU    | Ok     |
| 7   | P3479-07    | TS11                 | BF138906.D     | 10 Aug 2024 13:16 |                         | RC/JU    | Ok     |
| 8   | P3464-02MS  | TOPSOILMS            | BF138907.D     | 10 Aug 2024 13:47 | Internal Standard Fail  | RC/JU    | Ok,M   |
| 9   | P3464-02MSD | TOPSOILMSD           | BF138908.D     | 10 Aug 2024 14:17 | Internal Standard Fail, | RC/JU    | Ok,M   |
| 10  | P3415-04MS  | MLS-15-70-85MS       | BF138909.D     | 10 Aug 2024 14:48 |                         | RC/JU    | Ok     |
| 11  | P3415-05MSD | MLS-15-70-85MSD      | BF138910.D     | 10 Aug 2024 15:19 |                         | RC/JU    | Ok     |
| 12  | P3476-01    | MUL04                | BF138911.D     | 10 Aug 2024 15:49 |                         | RC/JU    | Ok     |
| 13  | P3466-04    | 932-K1-WP0-0.25-0802 | BF138912.D     | 10 Aug 2024 16:20 |                         | RC/JU    | Ok     |
| 14  | P3466-05    | 919-J-WP0-0.25-08022 | BF138913.D     | 10 Aug 2024 16:51 |                         | RC/JU    | Ok     |
| 15  | P3476-07    | MUL07                | BF138914.D     | 10 Aug 2024 17:21 | Internal Standard Fail  | RC/JU    | ReRun  |
| 16  | P3457-01    | 924-K1-WS-080224     | BF138915.D     | 10 Aug 2024 17:52 |                         | RC/JU    | Ok     |
| 17  | P3457-02    | 932-K1-WS-080224     | BF138916.D     | 10 Aug 2024 18:22 |                         | RC/JU    | Ok     |
| 18  | P3466-01    | 924-K1-WP0-0.25-0802 | BF138917.D     | 10 Aug 2024 18:53 |                         | RC/JU    | Ok     |

Instrument ID: BNA\_F

**Daily Analysis Runlog For Sequence/QC Batch ID # BF081024**

| Review By                | Jagrut   | Review On                                               | 8/12/2024 5:10:00 PM |                      |          |
|--------------------------|----------|---------------------------------------------------------|----------------------|----------------------|----------|
| Supervise By             | mohammad | Supervise On                                            | 8/13/2024 5:17:21 AM |                      |          |
| SubDirectory             | BF081024 | HP Acquire Method                                       | BNA_F                | HP Processing Method | bf073024 |
| STD. NAME                |          | STD REF.#                                               |                      |                      |          |
| Tune/Reschk              |          | SP6573                                                  |                      |                      |          |
| Initial Calibration Stds |          | SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557 |                      |                      |          |
| CCC                      |          | SP6553                                                  |                      |                      |          |
| Internal Standard/PEM    |          | S12039,10ul/1000ul sample                               |                      |                      |          |
| ICV/I.BLK                |          | SP6559                                                  |                      |                      |          |
| Surrogate Standard       |          |                                                         |                      |                      |          |
| MS/MSD Standard          |          |                                                         |                      |                      |          |
| LCS Standard             |          |                                                         |                      |                      |          |

|    |             |                      |            |                   |                        |       |        |
|----|-------------|----------------------|------------|-------------------|------------------------|-------|--------|
| 19 | P3466-02MS  | 924-K1-WP0-0.25-0802 | BF138918.D | 10 Aug 2024 19:23 | MSD not ok             | RC/JU | Not Ok |
| 20 | P3466-03MSD | 924-K1-WP0-0.25-0802 | BF138919.D | 10 Aug 2024 19:54 | Internal Standard Fail | RC/JU | Not Ok |
| 21 | P3476-03    | MUL05                | BF138920.D | 10 Aug 2024 20:24 |                        | RC/JU | Ok     |
| 22 | P3476-05    | MUL06                | BF138921.D | 10 Aug 2024 20:54 | Internal Standard Fail | RC/JU | ReRun  |
| 23 | P3467-01    | 919-J-WS-080224      | BF138922.D | 10 Aug 2024 21:25 |                        | RC/JU | Ok     |

M : Manual Integration