

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3429 SAS No.: P3429 SDG NO.: P3429
Lab File ID: VN083246.D Date Analyzed: 08/13/2024
Instrument ID: MSVOA_N Time Analyzed: 10:18
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	178090	8.22	308255	9.10	271780	11.87
UPPER LIMIT	356180	8.724	616510	9.6	543560	12.365
LOWER LIMIT	89045	7.724	154128	8.6	135890	11.365
EPA SAMPLE NO.						
931-K1-WS-073124	136049	8.22	260779	9.10	272432	11.87
925-K1-WS-073124	131513	8.22	255536	9.10	257054	11.87
TB-01-073124	147065	8.22	284471	9.10	291862	11.87
VN0813WBL01	151482	8.22	289244	9.10	291555	11.87
VN0813WBS01	178353	8.22	306441	9.10	263797	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.