

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3429 SAS No.: P3429 SDG NO.: P3429
Lab File ID: VN083295.D Date Analyzed: 08/14/2024
Instrument ID: MSVOA_N Time Analyzed: 12:41
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	165534	8.22	282606	9.10	250996	11.87
UPPER LIMIT	331068	8.724	565212	9.6	501992	12.365
LOWER LIMIT	82767	7.724	141303	8.6	125498	11.365
EPA SAMPLE NO.						
925-K1-WS-073124DL	130132	8.22	253353	9.10	260285	11.87
VN0814WBL01	145129	8.23	276647	9.10	286871	11.87
VN0814WBS01	164632	8.22	284346	9.10	250581	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.