



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: P3440

Project: Former Schlumberger Site Princeton NJ

RunNo.: LB131838

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	08/01/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.486	0.5	97	90-110	08/01/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.490	0.5	98	90-110	08/01/2024