

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: P3440
Project: Former Schlumberger Site Princeton NJ	Sample ID: P3440-01
Client ID: 923-K1-WS-080124DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifie	Duplicate Result	Conc. Qualifie	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		08/01/2024

Duplicate Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	P3440
Project:	Former Schlumberger Site Princeton NJ	Sample ID:	P3440-01
Client ID:	923-K1-WS-080124MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifie	Duplicate Result	Conc. Qualifie	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.93		0.92		2	0.86		08/01/2024