

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV72	Mercury	4.01	4.0	100	90 - 110	CV	08/13/2024	08:53	LB131960

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV35	Mercury	4.53	5.0	91	90 - 110	CV	08/13/2024	08:58	LB131960
CCV36	Mercury	4.77	5.0	95	90 - 110	CV	08/13/2024	09:33	LB131960
CCV37	Mercury	4.95	5.0	99	90 - 110	CV	08/13/2024	10:00	LB131960
CCV38	Mercury	4.86	5.0	97	90 - 110	CV	08/13/2024	10:30	LB131960
CCV39	Mercury	5.18	5.0	104	90 - 110	CV	08/13/2024	10:58	LB131960
CCV40	Mercury	5.09	5.0	102	90 - 110	CV	08/13/2024	11:24	LB131960

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440

Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2310	2500	92	90 - 110	P	08/07/2024	15:52	LB131900
	Antimony	937	1000	94	90 - 110	P	08/07/2024	15:52	LB131900
	Arsenic	903	1000	90	90 - 110	P	08/07/2024	15:52	LB131900
	Barium	489	520	94	90 - 110	P	08/07/2024	15:52	LB131900
	Beryllium	472	510	92	90 - 110	P	08/07/2024	15:52	LB131900
	Boron	2480	2500	99	90 - 110	P	08/07/2024	15:52	LB131900
	Cadmium	470	510	92	90 - 110	P	08/07/2024	15:52	LB131900
	Calcium	9780	10000	98	90 - 110	P	08/07/2024	15:52	LB131900
	Chromium	521	520	100	90 - 110	P	08/07/2024	15:52	LB131900
	Cobalt	487	520	94	90 - 110	P	08/07/2024	15:52	LB131900
	Copper	490	510	96	90 - 110	P	08/07/2024	15:52	LB131900
	Iron	10000	10000	100	90 - 110	P	08/07/2024	15:52	LB131900
	Lead	954	1000	95	90 - 110	P	08/07/2024	15:52	LB131900
	Magnesium	5520	6000	92	90 - 110	P	08/07/2024	15:52	LB131900
	Manganese	493	520	95	90 - 110	P	08/07/2024	15:52	LB131900
	Molybdenum	2710	2500	108	90 - 110	P	08/07/2024	15:52	LB131900
	Nickel	490	530	92	90 - 110	P	08/07/2024	15:52	LB131900
	Potassium	10300	9900	104	90 - 110	P	08/07/2024	15:52	LB131900
	Selenium	916	1000	92	90 - 110	P	08/07/2024	15:52	LB131900
	Silver	251	250	100	90 - 110	P	08/07/2024	15:52	LB131900
	Sodium	10500	10000	105	90 - 110	P	08/07/2024	15:52	LB131900
	Strontium	2500	2500	100	90 - 110	P	08/07/2024	15:52	LB131900
	Thallium	999	1000	100	90 - 110	P	08/07/2024	15:52	LB131900
	Tin	2630	2500	105	90 - 110	P	08/07/2024	15:52	LB131900
	Titanium	2640	2500	106	90 - 110	P	08/07/2024	15:52	LB131900
	Vanadium	475	500	95	90 - 110	P	08/07/2024	15:52	LB131900
	Zinc	1040	1000	104	90 - 110	P	08/07/2024	15:52	LB131900

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440

Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	102	100	102	80 - 120	P	08/07/2024	15:56	LB131900
	Antimony	51.2	50.0	102	80 - 120	P	08/07/2024	15:56	LB131900
	Arsenic	19.9	20.0	100	80 - 120	P	08/07/2024	15:56	LB131900
	Barium	101	100	101	80 - 120	P	08/07/2024	15:56	LB131900
	Beryllium	6.00	6.0	100	80 - 120	P	08/07/2024	15:56	LB131900
	Boron	102	100	102	80 - 120	P	08/07/2024	15:56	LB131900
	Cadmium	6.25	6.0	104	80 - 120	P	08/07/2024	15:56	LB131900
	Calcium	2040	2000	102	80 - 120	P	08/07/2024	15:56	LB131900
	Chromium	9.27	10.0	93	80 - 120	P	08/07/2024	15:56	LB131900
	Cobalt	29.6	30.0	99	80 - 120	P	08/07/2024	15:56	LB131900
	Copper	21.9	20.0	109	80 - 120	P	08/07/2024	15:56	LB131900
	Iron	99.5	100	100	80 - 120	P	08/07/2024	15:56	LB131900
	Lead	13.9	12.0	116	80 - 120	P	08/07/2024	15:56	LB131900
	Magnesium	2100	2000	105	80 - 120	P	08/07/2024	15:56	LB131900
	Manganese	20.6	20.0	103	80 - 120	P	08/07/2024	15:56	LB131900
	Molybdenum	201	200	101	80 - 120	P	08/07/2024	15:56	LB131900
	Nickel	40.0	40.0	100	80 - 120	P	08/07/2024	15:56	LB131900
	Potassium	1760	2000	88	80 - 120	P	08/07/2024	15:56	LB131900
	Selenium	18.3	20.0	91	80 - 120	P	08/07/2024	15:56	LB131900
	Silver	8.68	10.0	87	80 - 120	P	08/07/2024	15:56	LB131900
	Sodium	1730	2000	86	80 - 120	P	08/07/2024	15:56	LB131900
	Strontium	20.7	20.0	104	80 - 120	P	08/07/2024	15:56	LB131900
	Thallium	41.2	40.0	103	80 - 120	P	08/07/2024	15:56	LB131900
	Tin	36.0	40.0	90	80 - 120	P	08/07/2024	15:56	LB131900
	Titanium	40.8	40.0	102	80 - 120	P	08/07/2024	15:56	LB131900
	Vanadium	41.4	40.0	104	80 - 120	P	08/07/2024	15:56	LB131900
	Zinc	41.5	40.0	104	80 - 120	P	08/07/2024	15:56	LB131900

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440

Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9700	10000	97	90 - 110	P	08/07/2024	16:20	LB131900
	Antimony	4930	5000	99	90 - 110	P	08/07/2024	16:20	LB131900
	Arsenic	4910	5000	98	90 - 110	P	08/07/2024	16:20	LB131900
	Barium	10000	10000	100	90 - 110	P	08/07/2024	16:20	LB131900
	Beryllium	249	250	100	90 - 110	P	08/07/2024	16:20	LB131900
	Boron	4900	5000	98	90 - 110	P	08/07/2024	16:20	LB131900
	Cadmium	2490	2500	100	90 - 110	P	08/07/2024	16:20	LB131900
	Calcium	25000	25000	100	90 - 110	P	08/07/2024	16:20	LB131900
	Chromium	1010	1000	100	90 - 110	P	08/07/2024	16:20	LB131900
	Cobalt	2470	2500	99	90 - 110	P	08/07/2024	16:20	LB131900
	Copper	1230	1250	98	90 - 110	P	08/07/2024	16:20	LB131900
	Iron	4990	5000	100	90 - 110	P	08/07/2024	16:20	LB131900
	Lead	5060	5000	101	90 - 110	P	08/07/2024	16:20	LB131900
	Magnesium	24600	25000	98	90 - 110	P	08/07/2024	16:20	LB131900
	Manganese	2480	2500	99	90 - 110	P	08/07/2024	16:20	LB131900
	Molybdenum	5040	5000	101	90 - 110	P	08/07/2024	16:20	LB131900
	Nickel	2490	2500	100	90 - 110	P	08/07/2024	16:20	LB131900
	Potassium	24300	25000	97	90 - 110	P	08/07/2024	16:20	LB131900
	Selenium	4880	5000	98	90 - 110	P	08/07/2024	16:20	LB131900
	Silver	1240	1250	100	90 - 110	P	08/07/2024	16:20	LB131900
	Sodium	24000	25000	96	90 - 110	P	08/07/2024	16:20	LB131900
	Strontium	5000	5000	100	90 - 110	P	08/07/2024	16:20	LB131900
	Thallium	5160	5000	103	90 - 110	P	08/07/2024	16:20	LB131900
	Tin	5000	5000	100	90 - 110	P	08/07/2024	16:20	LB131900
	Titanium	4950	5000	99	90 - 110	P	08/07/2024	16:20	LB131900
	Vanadium	2460	2500	98	90 - 110	P	08/07/2024	16:20	LB131900
	Zinc	2540	2500	102	90 - 110	P	08/07/2024	16:20	LB131900
CCV02	Aluminum	9760	10000	98	90 - 110	P	08/07/2024	17:14	LB131900
	Antimony	4910	5000	98	90 - 110	P	08/07/2024	17:14	LB131900
	Arsenic	4900	5000	98	90 - 110	P	08/07/2024	17:14	LB131900
	Barium	10200	10000	102	90 - 110	P	08/07/2024	17:14	LB131900
	Beryllium	250	250	100	90 - 110	P	08/07/2024	17:14	LB131900
	Boron	4890	5000	98	90 - 110	P	08/07/2024	17:14	LB131900
	Cadmium	2520	2500	101	90 - 110	P	08/07/2024	17:14	LB131900

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440

Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Calcium	25500	25000	102	90 - 110	P	08/07/2024	17:14	LB131900
	Chromium	1010	1000	101	90 - 110	P	08/07/2024	17:14	LB131900
	Cobalt	2490	2500	100	90 - 110	P	08/07/2024	17:14	LB131900
	Copper	1240	1250	99	90 - 110	P	08/07/2024	17:14	LB131900
	Iron	5090	5000	102	90 - 110	P	08/07/2024	17:14	LB131900
	Lead	5100	5000	102	90 - 110	P	08/07/2024	17:14	LB131900
	Magnesium	24800	25000	99	90 - 110	P	08/07/2024	17:14	LB131900
	Manganese	2530	2500	101	90 - 110	P	08/07/2024	17:14	LB131900
	Molybdenum	5050	5000	101	90 - 110	P	08/07/2024	17:14	LB131900
	Nickel	2510	2500	100	90 - 110	P	08/07/2024	17:14	LB131900
	Potassium	24500	25000	98	90 - 110	P	08/07/2024	17:14	LB131900
	Selenium	4850	5000	97	90 - 110	P	08/07/2024	17:14	LB131900
	Silver	1250	1250	100	90 - 110	P	08/07/2024	17:14	LB131900
	Sodium	24200	25000	97	90 - 110	P	08/07/2024	17:14	LB131900
	Strontium	5100	5000	102	90 - 110	P	08/07/2024	17:14	LB131900
	Thallium	5130	5000	103	90 - 110	P	08/07/2024	17:14	LB131900
	Tin	5030	5000	100	90 - 110	P	08/07/2024	17:14	LB131900
	Titanium	5060	5000	101	90 - 110	P	08/07/2024	17:14	LB131900
	Vanadium	2490	2500	100	90 - 110	P	08/07/2024	17:14	LB131900
	Zinc	2560	2500	102	90 - 110	P	08/07/2024	17:14	LB131900
CCV03	Aluminum	9390	10000	94	90 - 110	P	08/07/2024	18:51	LB131900
	Antimony	4780	5000	96	90 - 110	P	08/07/2024	18:51	LB131900
	Arsenic	4720	5000	94	90 - 110	P	08/07/2024	18:51	LB131900
	Barium	9840	10000	98	90 - 110	P	08/07/2024	18:51	LB131900
	Beryllium	252	250	101	90 - 110	P	08/07/2024	18:51	LB131900
	Boron	4910	5000	98	90 - 110	P	08/07/2024	18:51	LB131900
	Cadmium	2440	2500	98	90 - 110	P	08/07/2024	18:51	LB131900
	Calcium	24700	25000	99	90 - 110	P	08/07/2024	18:51	LB131900
	Chromium	996	1000	100	90 - 110	P	08/07/2024	18:51	LB131900
	Cobalt	2430	2500	97	90 - 110	P	08/07/2024	18:51	LB131900
	Copper	1200	1250	96	90 - 110	P	08/07/2024	18:51	LB131900
	Iron	4920	5000	98	90 - 110	P	08/07/2024	18:51	LB131900
	Lead	4950	5000	99	90 - 110	P	08/07/2024	18:51	LB131900
	Magnesium	24200	25000	97	90 - 110	P	08/07/2024	18:51	LB131900

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440

Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Manganese	2460	2500	98	90 - 110	P	08/07/2024	18:51	LB131900
	Molybdenum	4970	5000	99	90 - 110	P	08/07/2024	18:51	LB131900
	Nickel	2440	2500	97	90 - 110	P	08/07/2024	18:51	LB131900
	Potassium	24200	25000	97	90 - 110	P	08/07/2024	18:51	LB131900
	Selenium	4670	5000	93	90 - 110	P	08/07/2024	18:51	LB131900
	Silver	1240	1250	99	90 - 110	P	08/07/2024	18:51	LB131900
	Sodium	23600	25000	94	90 - 110	P	08/07/2024	18:51	LB131900
	Strontium	4900	5000	98	90 - 110	P	08/07/2024	18:51	LB131900
	Thallium	4990	5000	100	90 - 110	P	08/07/2024	18:51	LB131900
	Tin	4860	5000	97	90 - 110	P	08/07/2024	18:51	LB131900
	Titanium	4890	5000	98	90 - 110	P	08/07/2024	18:51	LB131900
	Vanadium	2420	2500	97	90 - 110	P	08/07/2024	18:51	LB131900
	Zinc	2500	2500	100	90 - 110	P	08/07/2024	18:51	LB131900
CCV04	Aluminum	9190	10000	92	90 - 110	P	08/07/2024	19:48	LB131900
	Antimony	4780	5000	96	90 - 110	P	08/07/2024	19:48	LB131900
	Arsenic	4740	5000	95	90 - 110	P	08/07/2024	19:48	LB131900
	Barium	9880	10000	99	90 - 110	P	08/07/2024	19:48	LB131900
	Beryllium	242	250	97	90 - 110	P	08/07/2024	19:48	LB131900
	Boron	4700	5000	94	90 - 110	P	08/07/2024	19:48	LB131900
	Cadmium	2460	2500	98	90 - 110	P	08/07/2024	19:48	LB131900
	Calcium	24600	25000	98	90 - 110	P	08/07/2024	19:48	LB131900
	Chromium	995	1000	100	90 - 110	P	08/07/2024	19:48	LB131900
	Cobalt	2430	2500	97	90 - 110	P	08/07/2024	19:48	LB131900
	Copper	1200	1250	96	90 - 110	P	08/07/2024	19:48	LB131900
	Iron	5100	5000	102	90 - 110	P	08/07/2024	19:48	LB131900
	Lead	5000	5000	100	90 - 110	P	08/07/2024	19:48	LB131900
	Magnesium	23900	25000	96	90 - 110	P	08/07/2024	19:48	LB131900
	Manganese	2440	2500	98	90 - 110	P	08/07/2024	19:48	LB131900
	Molybdenum	4950	5000	99	90 - 110	P	08/07/2024	19:48	LB131900
	Nickel	2450	2500	98	90 - 110	P	08/07/2024	19:48	LB131900
	Potassium	24600	25000	98	90 - 110	P	08/07/2024	19:48	LB131900
	Selenium	4690	5000	94	90 - 110	P	08/07/2024	19:48	LB131900
	Silver	1240	1250	99	90 - 110	P	08/07/2024	19:48	LB131900
	Sodium	23900	25000	96	90 - 110	P	08/07/2024	19:48	LB131900

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3440
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Strontium	4890	5000	98	90 - 110	P	08/07/2024	19:48	LB131900
	Thallium	5020	5000	100	90 - 110	P	08/07/2024	19:48	LB131900
	Tin	4900	5000	98	90 - 110	P	08/07/2024	19:48	LB131900
	Titanium	4860	5000	97	90 - 110	P	08/07/2024	19:48	LB131900
	Vanadium	2410	2500	96	90 - 110	P	08/07/2024	19:48	LB131900
	Zinc	2500	2500	100	90 - 110	P	08/07/2024	19:48	LB131900