

**Metals**  
**- 5b -**  
**POST DIGEST SPIKE SUMMARY**

**Client:** JACOBS Engineering Group, Inc. **SDG No.:** P3440  
**Contract:** JACO05 **Lab Code:** CHEM **Case No.:** P3440 **SAS No.:** P3440  
**Matrix:** Water **Level:** LOW **Client ID:** 923-K1-WS-080124A  
**Sample ID:** P3440-01 **Spiked ID:** P3440-01A

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Silver  | ug/L  | 75 - 125               | 321              |   | 1.00             | U | 500            | 64            |      | P |