

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080524\
 Data File : BN033245.D
 Acq On : 05 Aug 2024 21:07
 Operator : MA/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Aug 05 23:12:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 22:59:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.474	152	1116	0.400	ng	-0.04
7) Naphthalene-d8	10.244	136	3936	0.400	ng	#-0.07
13) Acenaphthene-d10	14.101	164	2680	0.400	ng	-0.03
19) Phenanthrene-d10	16.890	188	5864	0.400	ng	-0.03
29) Chrysene-d12	21.104	240	6449	0.400	ng	-0.02
35) Perylene-d12	23.280	264	7359	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.228	112	4067	1.441	ng	-0.07
5) Phenol-d6	7.084	99	309	0.088	ng	0.09
8) Nitrobenzene-d5	8.728	82	5741	1.937	ng	-0.19
11) 2-Methylnaphthalene-d10	11.848	152	10073	1.673	ng	-0.08
14) 2,4,6-Tribromophenol	15.661	330	1656	1.446	ng	-0.06
15) 2-Fluorobiphenyl	12.732	172	15960	1.600	ng	-0.07
27) Fluoranthene-d10	18.929	212	27262	1.797	ng	-0.02
31) Terphenyl-d14	19.533	244	18805	1.423	ng	-0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.047	88	2463	1.782	ng	95
3) n-Nitrosodimethylamine	3.429	42	2236	1.913	ng	# 84
6) bis(2-Chloroethyl)ether	6.968	93	6447	2.200	ng	# 11
9) Naphthalene	10.287	128	18011	1.652	ng	# 86
10) Hexachlorobutadiene	10.500	225	3102	1.489	ng	# 98
12) 2-Methylnaphthalene	11.928	142	11647	1.635	ng	93
16) Acenaphthylene	13.834	152	18666	1.712	ng	99
17) Acenaphthene	14.165	154	13093	1.695	ng	99
18) Fluorene	15.170	166	17880	1.732	ng	100
21) 4-Bromophenyl-phenylether	16.071	248	5488	1.593	ng	96
22) Hexachlorobenzene	16.170	284	5490	1.423	ng	99
23) Atrazine	16.369	200	4801	1.892	ng	90
24) Pentachlorophenol	16.580	266	1962	1.368	ng	99
25) Phenanthrene	16.927	178	26931	1.680	ng	100
26) Anthracene	17.026	178	23929	1.814	ng	# 87
28) Fluoranthene	18.962	202	35521	1.832	ng	99
30) Pyrene	19.333	202	37401	1.420	ng	100
32) Benzo(a)anthracene	21.095	228	31989	1.569	ng	99
33) Chrysene	21.140	228	40041	1.652	ng	99
34) Bis(2-ethylhexyl)phtha...	21.006	149	1926	0.171	ng	95
36) Indeno(1,2,3-cd)pyrene	25.438	276	56857	1.960	ng	98
37) Benzo(b)fluoranthene	22.622	252	40817	1.592	ng	# 88
38) Benzo(k)fluoranthene	22.666	252	48357	1.651	ng	# 88
39) Benzo(a)pyrene	23.187	252	38767	1.677	ng	# 80
40) Dibenzo(a,h)anthracene	25.438	278	44999	1.964	ng	# 91
41) Benzo(g,h,i)perylene	26.113	276	50746	2.000	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080524\
 Data File : BN033245.D
 Acq On : 05 Aug 2024 21:07
 Operator : MA/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Aug 05 23:12:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 22:59:15 2024
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.474 min Scan# 61
 Delta R.T. -0.036 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 1116
 Ion Ratio Lower Upper
 152 100
 150 152.2 108.2 162.2
 115 82.1 70.5 105.7



#2
 1,4-Dioxane
 Concen: 1.782 ng
 RT: 3.047 min Scan# 23
 Delta R.T. 0.000 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07

Tgt Ion: 88 Resp: 2463
 Ion Ratio Lower Upper
 88 100
 43 26.7 18.6 28.0
 58 71.3 54.5 81.7









#3
n-Nitrosodimethylamine
Concen: 1.913 ng
RT: 3.429 min Scan# 70
Delta R.T. -0.029 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 2236
Ion Ratio Lower Upper
42 100
74 129.3 0.0 0.0#
44 5.9 9.6 14.4#



#4
2-Fluorophenol
Concen: 1.441 ng
RT: 5.228 min Scan# 325
Delta R.T. -0.065 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion: 112 Resp: 4067
Ion Ratio Lower Upper
112 100
64 60.9 18.0 27.0#
63 7.7 0.0 0.0#









#5
Phenol-d6
Concen: 0.088 ng
RT: 7.084 min Scan# 51
Delta R.T. 0.094 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 309
Ion Ratio Lower Upper
99 100
42 0.0 0.0 0.0
71 0.0 0.0 0.0



#6
bis(2-Chloroethyl)ether
Concen: 2.200 ng
RT: 6.968 min Scan# 566
Delta R.T. -0.087 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion: 93 Resp: 6447
Ion Ratio Lower Upper
93 100
63 35.9 1.1 1.7#
95 21.3 0.0 0.0#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.244 min Scan# 91
Delta R.T. -0.075 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	3936
Ion	Ratio	Lower	Upper
136	100		
137	14.3	14.1	21.1
54	11.9	13.8	20.8#
68	9.8	12.4	18.6#



#8
Nitrobenzene-d5
Concen: 1.937 ng
RT: 8.728 min Scan# 791
Delta R.T. -0.192 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:	82	Resp:	5741
Ion	Ratio	Lower	Upper
82	100		
128	41.4	54.8	82.2#
54	53.3	55.9	83.9#





#9

Naphthalene

Concen: 1.652 ng

RT: 10.287 min Scan# 91

Delta R.T. -0.096 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:128 Resp: 18011

Ion Ratio Lower Upper

128 100

129 11.9 15.4 23.0#

127 14.3 15.8 23.6#



#10

Hexachlorobutadiene

Concen: 1.489 ng

RT: 10.500 min Scan# 957

Delta R.T. -0.011 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07




Tgt Ion:225 Resp: 3102


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.2 50.2 75.4





#11

2-Methylnaphthalene-d10

Concen: 1.673 ng

RT: 11.848 min Scan# 1142

Delta R.T. -0.079 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07

Instrument :

BNA_N

ClientSampleId :





SSTDICC1.6

Tgt Ion:152 Resp: 10073

Ion Ratio Lower Upper

152 100

151 20.8 11.9 17.9#



#12

2-Methylnaphthalene

Concen: 1.635 ng

RT: 11.928 min Scan# 1163

Delta R.T. -0.075 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07




Tgt Ion:142 Resp: 11647


Ion Ratio Lower Upper

142 100

141 88.5 68.7 103.1

115 43.0 41.6 62.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.101 min Scan# 14
Delta R.T. -0.025 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 2680
Ion Ratio Lower Upper
164 100
162 105.1 84.6 127.0
160 51.5 42.4 63.6



#14
2,4,6-Tribromophenol
Concen: 1.446 ng
RT: 15.661 min Scan# 1626
Delta R.T. -0.055 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:330 Resp: 1656
Ion Ratio Lower Upper
330 100
332 95.1 73.7 110.5
141 55.3 44.8 67.2









#15
2-Fluorobiphenyl
Concen: 1.600 ng
RT: 12.732 min Scan# 11
Delta R.T. -0.068 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 15960
Ion Ratio Lower Upper
172 100
171 35.5 32.3 48.5
170 24.6 23.4 35.2



#16
Acenaphthylene
Concen: 1.712 ng
RT: 13.834 min Scan# 1457
Delta R.T. -0.036 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:152 Resp: 18666
Ion Ratio Lower Upper
152 100
151 19.3 16.1 24.1
153 12.7 10.2 15.2









#17
Acenaphthene
Concen: 1.695 ng
RT: 14.165 min Scan# 1492
Delta R.T. -0.036 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 13093
Ion Ratio Lower Upper
154 100
153 115.0 92.7 139.1
152 56.8 44.0 66.0



#18
Fluorene
Concen: 1.732 ng
RT: 15.170 min Scan# 1582
Delta R.T. -0.046 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:166 Resp: 17880
Ion Ratio Lower Upper
166 100
165 99.7 79.6 119.4
167 13.3 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.890 min Scan# 11
Delta R.T. -0.030 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 5864
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.4 10.2 15.4



#21
4-Bromophenyl-phenylether
Concen: 1.593 ng
RT: 16.071 min Scan# 1659
Delta R.T. -0.043 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:248 Resp: 5488
Ion Ratio Lower Upper
248 100
250 96.5 80.0 120.0
141 92.1 71.0 106.6









#22
Hexachlorobenzene
Concen: 1.423 ng
RT: 16.170 min Scan# 1667
Delta R.T. -0.030 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:284 Resp: 5490
Ion Ratio Lower Upper
284 100
142 49.1 38.8 58.2
249 34.1 27.8 41.6



#23
Atrazine
Concen: 1.892 ng
RT: 16.369 min Scan# 1683
Delta R.T. -0.030 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:200 Resp: 4801
Ion Ratio Lower Upper
200 100
173 30.7 29.2 43.8
215 44.9 41.5 62.3





#24

Pentachlorophenol

Concen: 1.368 ng

RT: 16.580 min Scan# 1

Delta R.T. -0.042 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:266 Resp: 1962

Ion Ratio Lower Upper

266 100

264 60.0 49.3 73.9

268 62.4 50.2 75.2



#25

Phenanthrene

Concen: 1.680 ng

RT: 16.927 min Scan# 1728

Delta R.T. -0.030 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07




Tgt Ion:178 Resp: 26931


Ion Ratio Lower Upper

178 100

176 19.5 15.7 23.5

179 15.3 12.4 18.6









#26
 Anthracene
 Concen: 1.814 ng
 RT: 17.026 min Scan# 11
 Delta R.T. -0.043 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:178 Resp: 23929
 Ion Ratio Lower Upper
 178 100
 176 18.8 10.5 15.7#
 179 15.0 8.6 12.8#



#27
 Fluoranthene-d10
 Concen: 1.797 ng
 RT: 18.929 min Scan# 1945
 Delta R.T. -0.021 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07

Tgt Ion:212 Resp: 27262
 Ion Ratio Lower Upper
 212 100
 106 14.2 11.4 17.0
 104 8.7 7.1 10.7









#28
Fluoranthene
Concen: 1.832 ng
RT: 18.962 min Scan# 1952
Delta R.T. -0.021 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:202 Resp: 35521
Ion Ratio Lower Upper
202 100
101 11.2 8.6 13.0
203 17.2 13.8 20.8



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.104 min Scan# 2403
Delta R.T. -0.021 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:240 Resp: 6449
Ion Ratio Lower Upper
240 100
120 17.4 13.3 19.9
236 29.7 23.8 35.6









#30
 Pyrene
 Concen: 1.420 ng
 RT: 19.333 min Scan# 2075
 Delta R.T. -0.016 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:202 Resp: 37401
 Ion Ratio Lower Upper
 202 100
 200 21.2 17.0 25.4
 203 17.9 14.4 21.6



#31
 Terphenyl-d14
 Concen: 1.423 ng
 RT: 19.533 min Scan# 2075
 Delta R.T. -0.016 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07

Tgt Ion:244 Resp: 18805
 Ion Ratio Lower Upper
 244 100
 212 10.1 9.9 14.9
 122 14.6 14.1 21.1





#32

Benzo(a)anthracene

Concen: 1.569 ng

RT: 21.095 min Scan# 2402

Delta R.T. -0.021 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 31989

Ion Ratio Lower Upper

228 100

226 27.4 22.4 33.6

229 19.9 16.6 25.0



#33

Chrysene

Concen: 1.652 ng

RT: 21.140 min Scan# 2407

Delta R.T. -0.020 min

Lab File: BN033245.D

Acq: 05 Aug 2024 21:07




Tgt Ion:228 Resp: 40041


Ion Ratio Lower Upper

228 100

226 31.0 25.1 37.7

229 19.6 16.2 24.2









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.171 ng
 RT: 21.006 min Scan# 21
 Delta R.T. -0.002 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:149 Resp: 1926
 Ion Ratio Lower Upper
 149 100
 167 27.4 19.8 29.6
 279 11.4 8.2 12.4



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.280 min Scan# 2846
 Delta R.T. -0.011 min
 Lab File: BN033245.D
 Acq: 05 Aug 2024 21:07

Tgt Ion:264 Resp: 7359
 Ion Ratio Lower Upper
 264 100
 260 26.9 21.4 32.2
 265 91.6 66.6 99.8









#36
Indeno(1,2,3-cd)pyrene
Concen: 1.960 ng
RT: 25.438 min Scan# 31
Delta R.T. -0.037 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:276 Resp: 56857
Ion Ratio Lower Upper
276 100
138 26.9 20.4 30.6
277 24.7 19.9 29.9



#37
Benzo(b)fluoranthene
Concen: 1.592 ng
RT: 22.622 min Scan# 2621
Delta R.T. -0.023 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07


Tgt Ion:252 Resp: 40817
Ion Ratio Lower Upper
252 100
253 23.3 23.3 34.9#
125 12.4 14.6 22.0#







#38
Benzo(k)fluoranthene
Concen: 1.651 ng
RT: 22.666 min Scan# 20
Delta R.T. -0.020 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 48357
Ion Ratio Lower Upper
252 100
253 23.6 23.7 35.5#
125 12.7 15.4 23.0#



#39
Benzo(a)pyrene
Concen: 1.677 ng
RT: 23.187 min Scan# 2814
Delta R.T. -0.022 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:252 Resp: 38767
Ion Ratio Lower Upper
252 100
253 24.5 28.4 42.6#
125 14.5 20.1 30.1#









#40
Dibenzo(a,h)anthracene
Concen: 1.964 ng
RT: 25.438 min Scan# 31
Delta R.T. -0.034 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:278 Resp: 44999
Ion Ratio Lower Upper
278 100
139 18.3 16.7 25.1
279 24.9 25.0 37.6#



#41
Benzo(g,h,i)perylene
Concen: 2.000 ng
RT: 26.113 min Scan# 3815
Delta R.T. -0.034 min
Lab File: BN033245.D
Acq: 05 Aug 2024 21:07

Tgt Ion:276 Resp: 50746
Ion Ratio Lower Upper
276 100
277 24.0 21.0 31.4
138 24.2 21.4 32.0

