

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3440 SAS No.: P3440 SDG NO.: P3440

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 08/06/2024

Lab File ID: BN033251.D Time Analyzed: 01:28

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	866	7.51	3240	10.34	2331	14.14
UPPER LIMIT	1732	8.01	6480	10.84	4662	14.637
LOWER LIMIT	433	7.01	1620	9.84	1165.5	13.637
EPA SAMPLE NO.						
01 PB162464BL	1306	7.63	4124	10.51	2380	14.21

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.