

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3440 SDG NO.: P3440

Lab File ID: BN033250.D

DFTPP Injection Date: 08/06/2024

Instrument ID: BNA_N

DFTPP Injection Time: 00:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	40.7
68	Less than 2.0% of mass 69	0.8 (1.8) 1
69	Mass 69 relative abundance	45.2
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	54.4
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26.7
365	Greater than 1% of mass 198	4.4
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	58.6
443	15.0 - 24.0% of mass 442	10.9 (18.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN033251.D	08/06/2024	01:28
PB162464BL	PB162464BL	BN033252.D	08/06/2024	02:04