

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
 Data File : BN033228.D
 Acq On : 03 Aug 2024 07:59
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 05 03:25:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 02 22:49:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.496	0.550	-10.9	76	0.00
3	n-Nitrosodimethylamine	0.419	0.488	-16.5	79	0.00
4 S	2-Fluorophenol	1.012	0.980	3.2	64	0.00
5 S	Phenol-d6	1.256	1.220	2.9	66	-0.02
6	bis(2-Chloroethyl)ether	1.050	0.885	15.7	57	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
8 S	Nitrobenzene-d5	0.301	0.293	2.7	78	0.00
9	Naphthalene	1.108	1.046	5.6	73	-0.01
10	Hexachlorobutadiene	0.212	0.199	6.1	72	-0.01
11 SURR	2-Methylnaphthalene-d10	0.612	0.548	10.5	71	0.00
12	2-Methylnaphthalene	0.724	0.605	16.4	67	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.171	0.112	34.5#	49#	0.00
15 S	2-Fluorobiphenyl	1.489	1.432	3.8	65	0.00
16	Acenaphthylene	1.627	1.592	2.2	72	0.00
17	Acenaphthene	1.153	1.123	2.6	67	0.00
18	Fluorene	1.541	1.413	8.3	63	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
20	4,6-Dinitro-2-methylphenol	0.056	0.013	76.8#	23#	0.00
21	4-Bromophenyl-phenylether	0.235	0.216	8.1	57	0.00
22	Hexachlorobenzene	0.263	0.229	12.9	54	0.00
23	Atrazine	0.173	0.161	6.9	63	0.00
24	Pentachlorophenol	0.098	0.053	45.9#	40#	0.00
25	Phenanthrene	1.093	0.994	9.1	57	0.00
26	Anthracene	0.900	0.809	10.1	62	0.00
27 SURR	Fluoranthene-d10	1.035	0.846	18.3	52	0.00
28	Fluoranthene	1.323	1.067	19.3	52	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
30	Pyrene	1.634	1.752	-7.2	53	0.00
31 S	Terphenyl-d14	0.820	0.856	-4.4	52	0.00
32	Benzo(a)anthracene	1.265	1.013	19.9	43#	0.00
33	Chrysene	1.504	1.600	-6.4	53	0.00
34	Bis(2-ethylhexyl)phthalate	0.700	0.763	-9.0	64	0.00
35 I	Perylene-d12	1.000	1.000	0.0	63	0.00
36	Indeno(1,2,3-cd)pyrene	1.577	1.448	8.2	60	0.00
37	Benzo(b)fluoranthene	1.393	1.112	20.2	54	0.00
38	Benzo(k)fluoranthene	1.592	1.356	14.8	54	0.00
39 C	Benzo(a)pyrene	1.256	1.090	13.2	57	0.00
40	Dibenzo(a,h)anthracene	1.245	1.136	8.8	59	0.00
41	Benzo(g,h,i)perylene	1.379	1.307	5.2	60	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4