

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3440 SAS No.: P3440 SDG NO.: P3440

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/08/2024

Lab File ID: BF138856.D Time Analyzed: 10:19

Instrument ID: BNA\_F GC Column: DB-UI ID: 0.18 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #  | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|-------|---------------------|--------|
| 12 HOUR STD    | 53967               | 6.839 | 212450              | 8.12  | 111283              | 9.88   |
| UPPER LIMIT    | 107934              | 7.339 | 424900              | 8.622 | 222566              | 10.375 |
| LOWER LIMIT    | 26983.5             | 6.339 | 106225              | 7.622 | 55641.5             | 9.375  |
| EPA SAMPLE NO. |                     |       |                     |       |                     |        |
| 01 PB162463BS  | 48969               | 6.84  | 204092              | 8.12  | 110072              | 9.88   |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.