

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3440

SDG NO.: P3440

Lab File ID: BF138679.D

DFTPP Injection Date: 07/30/2024

Instrument ID: BNA_F

DFTPP Injection Time: 12:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	49.4
68	Less than 2.0% of mass 69	0.8 (1.9) 1
69	Mass 69 relative abundance	41.1
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	51.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.9
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	12.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.8 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF138680.D	07/30/2024	12:54
SSTDICC005	SSTDICC005	BF138681.D	07/30/2024	13:25
SSTDICC010	SSTDICC010	BF138682.D	07/30/2024	13:56
SSTDICC020	SSTDICC020	BF138683.D	07/30/2024	14:25
SSTDICCC040	SSTDICCC040	BF138684.D	07/30/2024	14:56
SSTDICC050	SSTDICC050	BF138685.D	07/30/2024	15:27
SSTDICC060	SSTDICC060	BF138686.D	07/30/2024	15:58
SSTDICC080	SSTDICC080	BF138687.D	07/30/2024	16:29