

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3440 SAS No.: P3440 SDG NO.: P3440
 Lab File ID: VN083246.D Date Analyzed: 08/13/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:18
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	132143	13.794				
UPPER LIMIT	264286	14.294				
LOWER LIMIT	66071.5	13.294				
EPA SAMPLE NO.						
923-K1-WS-080124	131455	13.79				
923-K1-WS-080124MS	121293	13.79				
923-K1-WS-080124MSD	119452	13.79				
922-K1-WS-080124	111960	13.79				
TB-01-080124	127173	13.79				
VN0813WBL01	128023	13.79				
VN0813WBS01	120363	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.