

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bn080324

SequenceID : bn080324

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria	_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria	_____	_____✓_____	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____✓_____	_____
Compound #37 Failing Marginally Low side in the SSTDCCC(BN033212.D)	_____	_____✓_____	_____
Compound #37 Failing Marginally Low side in the SSTDCCC(BN033228.D)	_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
a. B/N Fraction	_____	_____✓_____	_____
d. Acid Fraction	_____	_____✓_____	_____

7. Surrogate Recoveries Meet Criteria	_____	_____	<u>✓</u>
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
a. B/N Fraction			
d. Acid Fraction			
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria	_____	<u>✓</u>	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
a. B/N Fraction	<u>The Recovery and RPD Fail for Some Compounds in P3440-02MS/03MSD,P3415-05MSD due to Matrix interference</u>		
d. Acid Fraction			
9. Internal Standard Area/Retention Time Shift Meet Criteria	_____	<u>✓</u>	_____
Comments:			
Internal Standard Fail in P3426-02			
10. Extraction Holding Time Met	_____	_____	<u>✓</u>
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met	_____	_____	<u>✓</u>
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bn080324

SequenceID : bn080324 NA NO YES