

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3451 SAS No.: P3451 SDG NO.: P3451

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/08/2024

Lab File ID: BF138856.D Time Analyzed: 10:19

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	53967	6.839	212450	8.12	111283	9.88
UPPER LIMIT	107934	7.339	424900	8.622	222566	10.375
LOWER LIMIT	26983.5	6.339	106225	7.622	55641.5	9.375
EPA SAMPLE NO.						
01 PB162463BS	48969	6.84	204092	8.12	110072	9.88

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.