

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3451 SAS No.: P3451 SDG NO.: P3451
 Lab File ID: VN083273.D Date Analyzed: 08/14/2024
 Instrument ID: MSVOA_N Time Analyzed: 02:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	173401	8.22	306155	9.10	269814	11.87
UPPER LIMIT	346802	8.724	612310	9.6	539628	12.365
LOWER LIMIT	86700.5	7.724	153078	8.6	134907	11.365
EPA SAMPLE NO.						
921-J-WS-080124	130924	8.22	257809	9.10	269897	11.87
TB-03-080124	138901	8.22	271745	9.10	276868	11.87
924-K1-WP0-0.25-080224MS	147545	8.22	267105	9.10	236618	11.87
924-K1-WP0-0.25-080224MSD	154428	8.23	282107	9.10	251921	11.87
VN0813WBL02	150257	8.22	292620	9.11	300089	11.87
VN0813WBS03	163931	8.22	287598	9.10	244933	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.