

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3451 SAS No.: P3451 SDG NO.: P3451
 Lab File ID: VN083273.D Date Analyzed: 08/14/2024
 Instrument ID: MSVOA_N Time Analyzed: 02:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	130108	13.788				
UPPER LIMIT	260216	14.288				
LOWER LIMIT	65054	13.288				
EPA SAMPLE NO.						
921-J-WS-080124	123166	13.79				
TB-03-080124	120328	13.79				
924-K1-WP0-0.25-080224MS	114165	13.79				
924-K1-WP0-0.25-080224MSD	119630	13.79				
VN0813WBL02	135334	13.79				
VN0813WBS03	114554	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.