

Prep Standard - Chemical Standard Summary

Order ID : P3457

Test : Hexavalent Chromium

Prepbatch ID :

Sequence ID/Qc Batch ID: LB131856,

Standard ID :

WP107791,WP108658,WP108659,WP108907,WP108973,WP108974,WP108975,WP108976,WP108977,WP108978,WP108979,WP108980,WP108981,

Chemical ID :

E3769,M5211,W2606,W2651,W2652,W2979,W3112,

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
126	5N sulfuric acid	WP107791	05/07/2024	10/24/2024	Niha Farheen Shaik	None	None	Iwona Zarych
								05/07/2024

FROM 140.00000ml of M5211 + 860.00000ml of W2606 = Final Quantity: 1.000 L

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1993	HEXAVALENTCHROMIUM STOCK STD 1, 50PPM	WP108658	07/09/2024	01/09/2025	Rubina Mughal	WETCHEM_SCALE_5 (WC	None	Iwona Zarych
						SC-5)		07/09/2024

FROM 0.14140gram of W2651 + 1000.00000ml of W3112 = Final Quantity: 1000.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1994	HEXAVALENTCHROMIUM STOCK STD 2, 50PPM	WP108659	07/09/2024	01/09/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 07/09/2024
<u>FROM</u> 0.14140gram of W2652 + 1000.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
114	hexavalent chromium color reagent	WP108907	07/30/2024	08/06/2024	Iwona Zarych	WETCHEM_S CALE_5 (WC SC-5)	None	Mohan Bera 08/02/2024
<u>FROM</u>	0.25000gram of W2979 + 50.00000ml of E3769 = Final Quantity: 50.000 ml							

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1103	HEX CHROME INTERMEDIATE STD SOURCE 1 (5PPM)	WP108973	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PL PETTE_3	Jignesh Parikh 08/02/2024
<p>FROM 9.00000ml of W3112 + 1.00000ml of WP108658 = Final Quantity: 10.000 ml (WC)</p>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
110	calibration std. hexchrome 0 ppm	WP108974	08/02/2024	08/03/2024	Iwona Zarych	None	None	Jignesh Parikh 08/02/2024
<p>FROM 100.00000ml of W3112 = Final Quantity: 100.000 ml</p>								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
109	calibration std. hexchrome 0.01 ppm	WP108975	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
FROM 99.80000ml of W3112 + 0.20000ml of WP108973 = Final Quantity: 100.000 ml (WC)								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3800	Calibration Std Hexachrome 0.025 ppm	WP108976	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
<u>FROM</u> 99.50000ml of W3112 + 0.50000ml of WP108973 = Final Quantity: 100.000 ml <div style="text-align: right;">(WC)</div>								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
108	Calibration Std. hexchrome 0.05 ppm	WP108977	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
FROM 99.00000ml of W3112 + 1.00000ml of WP108973 = Final Quantity: 100.000 ml (WC)								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
107	Calibration Std. hexchrome 0.1 ppm	WP108978	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
FROM 99.80000ml of W3112 + 0.20000ml of WP108658 = Final Quantity: 100.000 ml (WC)								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3808	Calibration and CCV std HexChrome 0.5PPM	WP108979	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
<u>FROM</u> 99.00000ml of W3112 + 1.00000ml of WP108658 = Final Quantity: 100.000 ml <div></div>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3809	Calibration std HexChrome 1.0PPM	WP108980	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
FROM 98.00000ml of W3112 + 2.00000ml of WP108658 = Final Quantity: 100.000 ml (WC)								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3804	Hexavalent Chromium ICV-LCS Std	WP108981	08/02/2024	08/03/2024	Iwona Zarych	None	WETCHEM_PIPETTE_3	Jignesh Parikh 08/02/2024
<u>FROM</u> 99.00000ml of W3112 + 1.00000ml of WP108659 = Final Quantity: 100.000 ml <div style="text-align: right;">(WC)</div>								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	01/12/2025	07/12/2024 / Rajesh	07/02/2024 / Rajesh	E3769

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	22D0862014	01/20/2025	08/22/2022 / mohan	04/26/2022 / mohan	M5211

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	10/24/2024	10/24/2019 / apatel	10/24/2019 / apatel	W2606

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AA13450-36 / Potassium Dichromate, 500g(NEW)	T15F019	01/24/2030	01/24/2020 / apatel	01/24/2020 / apatel	W2651

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	P188-500 / Potassium Dichromate, 500g(new-2nd lot)	194664	01/24/2030	01/24/2020 / apatel	01/24/2020 / apatel	W2652

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	31390 / 1,5-Diphenylcarbazine	MKCR6636	12/09/2027	12/09/2022 / lwona	12/09/2022 / lwona	W2979

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Product No.: 13450
Product: Potassium dichromate, ACS, 99.0% min
Lot No.: T15F019

Test	Limits	Results
Appearance	Orange-red crystals	Orange-red crystals
Identification	To Pass	Passes
Purity	99.0 % min	99.67 %
Insoluble matter	0.005 % max	0.004 %
Loss on drying	0.05 % max	0.03 %
Chloride	0.001 % max	< 0.001 %
Sulfate	0.005 % max	< 0.005 %
Iron	0.001 % max	< 0.001 %
Calcium	0.003 % max	0.0012 %
Sodium	0.02 % max	0.0047 %

Order our products online [alfa.com](https://www.thermofisher.com/alfa)

This document has been electronically generated and does not require a signature.

This is to certify that units of the lot number above were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The above information is the actual analytical results obtained.



Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	P188	Quality Test / Release Date	08/12/2019
Lot Number	194664		
Description	POTASSIUM DICHROMATE, A.C.S.		
Country of Origin	United States	Suggested Retest Date	Aug/2024
Chemical Origin	Inorganic-non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		
Chemical Comment			

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Fine, orange-red crystals
ASSAY	%	>= 99	99.2
CALCIUM	%	<= 0.003	<0.003
CHLORIDE	%	<= 0.001	<0.001
LOSS ON DRYING @ 105 C	%	<= 0.05	<0.05
SULFATE (SO4)	%	<= 0.005	<0.005
INSOLUBLE MATTER	%	<= 0.005	0.003
IRON (Fe)	%	<= 0.001	<0.001
SODIUM (Na)	%	<= 0.02	<0.02
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST

Jerusa Bailey-Wyche

Quality Assurance Specialist - Certificate of Analysis Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent

For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 23H1462005

Manufactured Date: 2023-07-26

Expiration Date: 2026-07-25

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 7/21/24

E 3769

Ken Koehnlein
Sr. Manager, Quality Assurance

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

avantor™



Material No.: 9673-33

Batch No.: 22D0862014

Manufactured Date: 2022-02-23

Retest Date: 2027-02-22

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.5 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Substances Reducing Permanganate (as SO ₂)	≤ 2 ppm	< 2 ppm
Ammonium (NH ₄)	≤ 1 ppm	< 1 ppm
Chloride (Cl)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO ₃)	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO ₄)	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities – Aluminum (Al)	≤ 30.0 ppb	1.7 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities – Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities – Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gold (Au)	≤ 10.0 ppb	< 0.2 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities – Iron (Fe)	≤ 50.0 ppb	2.0 ppb
Trace Impurities – Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities – Magnesium (Mg)	≤ 7.0 ppb	0.6 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities – Nickel (Ni)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities – Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se)	≤ 50.0 ppb	12.1 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	4.4 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium



Material No.: 9673-33
Batch No.: 22D0862014

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	6.2 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.6 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.
Jamie Ethier
Vice President Global Quality

W 2979

Rec: 12/09/22

exp. 12/09/27

Product Name:

1,5-Diphenylcarbazide - ACS reagent

Product Number:

259225

Batch Number:

MKCR6636

Brand:

SIAL

CAS Number:

140-22-7

MDL Number:

MFCD00003013

Formula:


C₁₃H₁₄N₄O

Formula Weight:

242.28 g/mol


Quality Release Date:

02 JUN 2022



Certificate of Analysis

Test	Specification	Result
Appearance (Color)	Conforms to Requirements	Pink
Off-White to Pink, Light Purple or Tan		
Appearance (Form)	Powder or Chunks	Powder
Melting Point	173.0 - 176.0 °C	173.0 °C
Infrared Spectrum	Conforms to Structure	Conforms
Residue on ignition (Ash)	≤ 0.05 %	0.01 %
15 minutes, 800 Degrees Celsius		
Solubility	Pass	Pass
Sensitivity Test	Pass	Pass
Meets ACS Requirements	Current ACS Specification	Conforms



Larry Coers, Director
Quality Control
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

