

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: P3457

Project: Former Schlumberger Site Princeton NJ

Run No.: LB131856

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	lb131856BS							
Hexavalent Chromium	mg/L	0.5	0.50		100	1	90-111	08/02/2024