

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: P3457

Contract: JACO05

Lab Code: CHEM

Case No.: P3457

SAS No.: P3457

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB162680BS Mercury	ug/L	4.0	3.62		90	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: P3457

Contract: JACO05

Lab Code: CHEM

Case No.: P3457

SAS No.: P3457

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB162950BS							
Aluminum	ug/L	10000	10500		105	80 - 120	P
Antimony	ug/L	500	499		100	80 - 120	P
Arsenic	ug/L	500	506		101	80 - 120	P
Barium	ug/L	2500	2510		100	80 - 120	P
Beryllium	ug/L	500	540		108	80 - 120	P
Cadmium	ug/L	500	500		100	80 - 120	P
Calcium	ug/L	50000	54500		109	80 - 120	P
Chromium	ug/L	500	516		103	80 - 120	P
Cobalt	ug/L	500	519		104	80 - 120	P
Copper	ug/L	5000	5230		105	80 - 120	P
Iron	ug/L	25000	28600		114	80 - 120	P
Lead	ug/L	2500	2530		101	80 - 120	P
Magnesium	ug/L	50000	53000		106	80 - 120	P
Manganese	ug/L	5000	5120		102	80 - 120	P
Nickel	ug/L	500	519		104	80 - 120	P
Potassium	ug/L	25000	25900		104	80 - 120	P
Selenium	ug/L	500	510		102	80 - 120	P
Silver	ug/L	500	518		104	80 - 120	P
Sodium	ug/L	50000	53800		108	80 - 120	P
Thallium	ug/L	500	508		102	80 - 120	P
Vanadium	ug/L	500	513		103	80 - 120	P
Zinc	ug/L	5000	5260		105	80 - 120	P