

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG No.: P3457

Instrument ID: MSVOA_X Calibration Date/Time: 08/09/2024 11:01

Lab File ID: VX042986.D Init. Calib. Date(s): 08/06/2024 08/06/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 18:18 20:15

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	51.10	-2.206001	20