

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG NO.: P3457
 Lab File ID: VN083197.D Date Analyzed: 08/10/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	189834	8.22	320853	9.10	279656	11.87
UPPER LIMIT	379668	8.724	641706	9.6	559312	12.365
LOWER LIMIT	94917	7.724	160427	8.6	139828	11.365
EPA SAMPLE NO.						
932-K1-WS-080224	139998	8.23	267123	9.10	272147	11.87
VN0810WBL02	124598	8.23	239076	9.10	236621	11.87
VN0810WBS01	169456	8.22	295589	9.10	256697	11.87
VN0810WBSD01	160373	8.22	278670	9.10	245098	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.