

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG NO.: P3457
 Lab File ID: VX042963.D Date Analyzed: 08/09/2024
 Instrument ID: MSVOA_X Time Analyzed: 00:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	230896	5.55	388392	6.76	344983	10.06
UPPER LIMIT	461792	6.05	776784	7.263	689966	10.555
LOWER LIMIT	115448	5.05	194196	6.263	172492	9.555
EPA SAMPLE NO.						
TB-01-080224	158206	5.55	312445	6.76	293579	10.06
VX0808WBL02	150400	5.55	282725	6.76	268240	10.06
VX0808WBS02	216431	5.55	372967	6.76	328816	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.