

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG NO.: P3457
 Lab File ID: VX042986.D Date Analyzed: 08/09/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	176162	12.024				
UPPER LIMIT	352324	12.524				
LOWER LIMIT	88081	11.524				
EPA SAMPLE NO.						
924-K1-WS-080224	96376	12.02				
VX0809WBL01	119442	12.02				
VX0809WBS01	163980	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.