

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG NO.: P3457
Lab File ID: VX042986.D Date Analyzed: 08/09/2024
Instrument ID: MSVOA_X Time Analyzed: 11:01
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	230041	5.55	373855	6.76	344403	10.06
UPPER LIMIT	460082	6.05	747710	7.257	688806	10.555
LOWER LIMIT	115021	5.05	186928	6.257	172202	9.555
EPA SAMPLE NO.						
924-K1-WS-080224	148730	5.55	240279	6.76	218815	10.06
VX0809WBL01	186614	5.55	300238	6.76	269204	10.06
VX0809WBS01	232797	5.55	392597	6.76	347845	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.