

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG NO.: P3457
 Lab File ID: VN083197.D Date Analyzed: 08/10/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	135305	13.794				
UPPER LIMIT	270610	14.294				
LOWER LIMIT	67652.5	13.294				
EPA SAMPLE NO.						
932-K1-WS-080224	121838	13.79				
VN0810WBL02	101434	13.79				
VN0810WBS01	116986	13.79				
VN0810WBSD01	113742	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.