

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3457 SAS No.: P3457 SDG NO.: P3457
 Lab File ID: VX042963.D Date Analyzed: 08/09/2024
 Instrument ID: MSVOA_X Time Analyzed: 00:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	171900	12.018				
UPPER LIMIT	343800	12.518				
LOWER LIMIT	85950	11.518				
EPA SAMPLE NO.						
TB-01-080224	121629	12.02				
VX0808WBL02	109749	12.02				
VX0808WBS02	153912	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.