



# SAMPLE DATA

A

B

C

## Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/02/24 13:25
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/02/24
Client Sample ID:	919-J-WS-080224	SDG No.:	P3467
Lab Sample ID:	P3467-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Dissolved Hexavalent Chromium	0.0030	U	1	0.0030	0.010	mg/L		08/03/24 08:44	7196A

Comments:

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements  
 H = Sample Analysis Out Of Hold Time

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits



LAB CHRONICLE

OrderID:	P3467	OrderDate:	8/2/2024 4:30:00 PM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger Site Princeton NJ
Contact:	Mary I. Murphy	Location:	D21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P3467-01	919-J-WS-080224	WATER	Hexavalent Chromium	7196A	08/02/24 13:25		08/03/24 08:44	08/02/24