



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: P3467

Project: Former Schlumberger Site Princeton NJ

RunNo.: LB131866

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	08/03/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	08/03/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	08/03/2024